

Results: The volumes of the single-cycle ITVs vary up to 20% around their mean. Averaged over all four datasets, the median CTV V95% for the ITV00, ITV25, ITV50, and ITV75 are 98.7%, 97.6%, 97.2%, and 94.3% respectively. In the same order, median D5-D95% are 9.4%, 10.1%, 10.6%, and 12.0%. The median ipsilateral mean lung doses are 14.8%, 13.8%, 13.2%, and 12.4%.

Conclusion: Breathing variations can have a significant impact on the ITV definition. Our probabilistic ITV definition can effectively account for variable motion, resulting in acceptable target coverage and substantial dose reduction to the involved lung. The ITV50 approach provided the best compromise between the two clinical objectives.

Quality assurance for the adaptive workflow on the MRI linac

Type: Physics

Presenting Author:

Lotte Wilke

Authors:

L. Wilke, S. Ehrbar, M. Bogowicz, J. Kraysenbühl, N. Andratschke, M. Guckenberger, S. Tanadini-Lang

Institutions:

Universitätsspital Zürich

Department:

Klinik für Radio-Onkologie

Aims: The new MRIdian MR Linac enables online adaptive treatment for daily changing anatomy of the patient. For these modulated treatments, no measurement of the treatment plan can be performed prior to treatment. Besides performing a secondary calculation within the system, we set up a chain of additional tests to ensure a safe treatment.

Methods: For all patients the original treatment plan as well as all adapted plans were measured on an IMRT verification phantom (Delta 4, Scandidos). During the adaptive process, we did a point dose verification in an independent Software (Radcalc). Additionally, we calculated a value that is representative of the integral dose delivered to the patient and compared this value to the original plan. To ensure that the electron densities are correctly assigned to the new MRI, we compared the equivalent pathlength of the original and the adapted plan.

Results: For the first 10 Patients, all measurements of the original plan passed the gamma-analysis with a 3%/3 mm criterion of on average 99.9% (range 99.6–100%). The adapted plans had equally good passing rates (average 99.8%, range 97.2–100%). For 89% of the plans the independent point dose calculation agreed within 10% with the original plan one (average 5.9%, range 0.3–21.4%). Values above 10% were connected to a lung treatment, where the scatter is incorrectly taken into account in the point dose calculation. The value representing the integral dose also agreed within 10% between the original plan and the treated adapted plan (average 4.0%, range 0.26–9.45%). The effective pathlength differed less than 2 cm from the original plan except for one case, where the change was traced back to a change in anatomy.

Conclusion: We set up a QA chain for the online adaptive planning on the MRIdian MR Linac which ensures a save treatment of the patient.

Influence of magnetic field transport parameters on accuracy and efficiency of electron transport in EGSnrc

Type: Physics

Presenting Author:

Gian Guyer

Authors:

G. Guyer¹, R. Kueng¹, W. Volken¹, M. F. M. Stampanoni², P. Manser¹, M. K. Fix¹

Institutions:

¹Inselspital, Bern University Hospital and University of Bern, ²ETH Zürich and PSI, Villigen

Department:

¹Division of Medical Radiation Physics and Department of Radiation Oncology, ²Institute for Biomedical Engineering

Aims: In context of MR guided radiotherapy, accurate and efficient dose calculation and therefore describing the propagation of electrons in a magnetic field is essential. The aim is to quantify the influence of the EGSnrc magnetic field transport parameters on the electron transport in a magnetic field in a local spherical geometry and a macroscopic phantom.

Methods: The EGSnrc 2019 version provides two different macros for handling magnetic fields: the EMF macro and a novel enhanced EEMF macro. Both are applied to calculate the deflection angle of 15 MeV electrons in a sphere (3 mm radius) due to a perpendicular 1.5 T magnetic field in vacuum and water with varying maximal electron step size. The deflection angles are compared to theoretical calculations. Next, the same parameters are used to calculate the dose in a water phantom with a 1 cm air slab in 1 cm depth for a 15 MeV electron pencil beam with a magnetic field of 1.5 T perpendicular to the incident beam direction. The resulting dose distributions are investigated by comparing relative dose differences in the air slab.

Results: For the EMF, varying the step size shows substantial differences in the deflection angle with an error up to 19% compared to the theoretical calculation. The EEMF shows only differences of up to 0.2%. The EMF shows dose differences of up to 6% in air but only 0.3% for a small step size. For the EEMF, the dose differences in air are up to 0.2% for all step sizes. The EMF computation time is 30% faster than the EEMF for the same step size but increases exponentially with decreasing step size.

Conclusion: The novel EEMF shows more accurate results for electron transport in a magnetic field when compared to EMF with the same step size. Although EMF can yield as accurate results as EEMF with a small step size, due to efficiency, EEMF should be preferred over EMF.

Dynamic trajectory radiotherapy versus HyperArc: treatment plan comparisons for tumors in the brain

Type: Physics

Presenting Author:

Peter Manser

Authors:

P. Manser, S. Mueller, D. Frei, W. Volken, D. Terribilini, D. Frauchiger, A. Joosten, D. Henzen, E. Herrmann, D. M. Aebersold, M. K. Fix

Institutions:

Inselspital, Bern University Hospital, and University of Bern

Department:

Division of Medical Radiation Physics

Purpose: HyperArc applies a pre-defined set of non-coplanar arcs for treatments of brain tumors, while dynamic trajectory radiotherapy (DTRT) additionally includes dynamic couch and collimator rotations during beam on to increase degrees of freedom for the selection of beam directions. The aim of this work is to investigate the potential benefit of DTRT compared to HyperArc for brain tumors.

Materials & methods: Five clinically motivated brain cases are included in this study. The target volumes range from 3 to 270 cm³ and are located at different areas in the brain with according organs at risk (OARs). A framework utilizing the Eclipse Scripting Research API was used to determine gantry-table and gantry-collimator paths based on contoured structures. These paths served as input for the multi-leaf sequence optimization using a research version of the VMAT optimization algorithm in Eclipse. Additionally, a HyperArc treatment plan was generated for each of the five cases. Resulting dose distributions for DTRT and HyperArc plans were compared based on DVH parameters. The deliverability of DTRT dose distributions was confirmed by gafchromic film measurements.