

## Article

# A Survey on the Use of Spirometry in Small Animal Anaesthesia and Critical Care

Mathieu Raillard <sup>1,2,3,4,\*</sup>, Olivier Levionnois <sup>3</sup>  and Martina Mosing <sup>4</sup> 

<sup>1</sup> AniCura Regiondjursjukhuset Bagarmossen, Ljusnevägen 17, 128 48 Bagarmossen, Sweden

<sup>2</sup> School of Veterinary Science, Faculty of Science, The University of Sydney, Evelyn Williams Building No. B10, Sydney, NSW 2006, Australia

<sup>3</sup> Anaesthesiology Section, Department of Clinical Veterinary Sciences, Vetsuisse Faculty, University of Berne, 3012 Berne, Switzerland; olivier.levionnois@vetsuisse.unibe.ch

<sup>4</sup> School of Veterinary and Life Sciences, College of Veterinary Medicine, Murdoch University, 90 South Street, Murdoch, WA 6150, Australia; M.Mosing@murdoch.edu.au

\* Correspondence: mathieu\_raillard@yahoo.it

**Simple Summary:** Spirometry can be used to monitor airway pressures, flows, and volumes. Its relevance in small animal anaesthesia is documented. The way the Veterinary Anaesthesia and Intensive Care communities use spirometry was not found in the literature. The objective was to document the use of spirometry and ventilation settings in small animal anaesthesia and intensive care through a descriptive, open, online, anonymous survey. The survey was advertised on social media and via email. Participation was voluntary. The google forms platform was used. It consisted of eight sections in English. Simple, descriptive analyses were applied. There were 128 respondents. Respondents used spirometry more when dogs were mechanically ventilated as opposed to spontaneously breathing. Over 3/4 of the respondents considered spirometry essential in “selected” (43%) or “most” cases (33%). Multiple devices and technologies were used. The majority of the respondents were not directly involved in or informed about the calibration of their equipment. More information on variables monitored and technical background on spirometers is required.

**Abstract:** The objective was to document the use of spirometry and ventilation settings in small animal anaesthesia and intensive care through a descriptive, open, online, anonymous survey. The survey was advertised on social media and via email. Participation was voluntary. The google forms platform was used. It consisted of eight sections in English: demographic information, use of spirometry in spontaneously ventilating/mechanically ventilated dogs, need for spirometry, equipment available and calibration status, ventilation modes, spirometry displays, compliance ( $C_{RS}$ ) and resistance ( $R_{RS}$ ) of the respiratory system. Simple descriptive analyses were applied. There were 128 respondents. Respondents used spirometry more in ventilated dogs than during spontaneous breathing. Over 3/4 of the respondents considered spirometry essential in “selected” (43%) or “most” cases (33%). Multiple devices and technologies were used. The majority of the respondents were not directly involved in or informed about the calibration of their equipment. Of all displays, pressure-volume loops were the most common. Values of  $C_{RS}$  and  $R_{RS}$  were specifically monitored in more than 50% of cases by 44% of the respondents only. A variety of ventilation modes was used. Intensivists tend to use smaller VT than anaesthetists. More information on reference intervals of  $C_{RS}$  and  $R_{RS}$  and technical background on spirometers is required

**Keywords:** compliance; resistance; spirometry; survey; tidal volume; ventilation



**Citation:** Raillard, M.; Levionnois, O.; Mosing, M. A Survey on the Use of Spirometry in Small Animal Anaesthesia and Critical Care. *Animals* **2022**, *12*, 239. <https://doi.org/10.3390/ani12030239>

Academic Editor:  
Nathalie Kirschvink

Received: 14 October 2021

Accepted: 15 January 2022

Published: 19 January 2022

**Publisher's Note:** MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



**Copyright:** © 2022 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

## 1. Introduction

Spirometry can be used to monitor continuously airway pressures, flows, and volumes and calculate compliance ( $C_{RS}$ ) and resistance ( $R_{RS}$ ) of the respiratory system. Its relevance in small animal anaesthesia is documented [1]. The way the Veterinary Anaesthesia and

Intensive Care communities use spirometry was not found in the literature. We investigated this through an online survey.

We had several goals. Firstly, we wanted to gather some habits of the community on the use of and perceived need for spirometry. Secondly, we wanted to obtain an overview of the equipment available and its calibration status. Thirdly, we wanted to investigate what spirometry displays were frequently monitored. Fourthly, we wanted to find out whether participants specifically considered  $C_{RS}$  and  $R_{RS}$ . Given the effects of ventilation modes on  $C_{RS}$  [2] and  $R_{RS}$  in anaesthetized dogs, information on commonly used ventilation settings was also sought. The parts of the survey that were not equipment-related (by nature, non-species specific) focused on dogs.

## 2. Materials and Methods

### 2.1. Survey Design

The CHERRIES (Checklist for Reporting Results of Internet E-Surveys) statement was followed [3]. This was a descriptive, open, anonymous survey. Convenience sampling (a method of selection from a conveniently available pool of respondents) was used.

### 2.2. Institutional Review Board Approval

Ethics were managed under the requirements of the University of Bern, where the research was undertaken. The study was conducted according to the guidelines of the Declaration of Helsinki. Ethical approval was waived for this study after review from the local ethical committee (Gesundheits-, Sozial- und Integrationsdirektion Kantonale Ethikkommission für die Forschung, Kanton Bern) because this project was not subject to any authorization according to national law (Humanforschungsgesetz, Art. 2, Abs 1).

### 2.3. Informed Consent

The introduction page informed the participants on the objectives, number of questions, and expected time needed to answer the survey. It also reported: "The data collected will remain confidential. No information to identify you or your practice will be collected. (...) You are undertaking this survey on a voluntary basis. You can decide at any time that you no longer want to participate. (...) By clicking "NEXT" you certify that you have read and understood the purpose of this study, your rights as an informed participant and that your answers will be used for analysis and publication."

### 2.4. Development and Pretesting

The survey was developed to be comprehensive and easily filled. It was sent to two ECVAA diplomates (European College of Veterinary Anaesthesia and Analgesia) with interest in respiratory mechanics, and one biomedical engineer with expertise in veterinary anaesthesia. The Survey was altered following their feedback (unclear content was reformulated, some questions were added/removed).

### 2.5. Recruitment Process and Description of the Target Population

The survey was advertised on social media (Facebook, LinkedIn, and Twitter) on the pages of the AVA (Association of Veterinary Anaesthetists), SEAAV (Sociedad Española de Anestesia y Analgesia Veterinaria), Anaesthesia and Emergency and Critical care chapters of the ANZCVS (Australia and New Zealand College of Veterinary Scientists), and EVECCS (European Veterinary Emergency and Critical Care Society). Links were sent using a shared email list of the international Veterinary Anaesthesia and Analgesia community ("acvalist"), the EVECCS newsletter, and through the secretaries of the ANZCVS Anaesthesia and Emergency and Critical care chapters and the ECVAA. Other institutions (including the American College of Emergency and Critical Care, ACVECC) were also contacted. A selection of survey announcements is presented in Appendix A.

## 2.6. Survey Administration

Google forms was used to host the survey. Participation was voluntary; no incentives were offered. The survey was conducted between February and July 2018. There were eight sections, composed of two to eight items each (39 multiple choice or open-ended questions, all in English). Items were not randomized. Adaptive questioning was not used, although some questions were facultative. Respondents could review and change their answers through the “Back” button until they submitted the full questionnaire.

No effort was made to prevent multiple entries from the same individual (no IP check or specific cookies used). Responses were received only if the respondents completed the final screen of the survey and submitted their questionnaire. View rate, participation, and completion rates were not calculated.

## 2.7. Questionnaire

Section one asked demographic information (three questions), including background, working environment, and country. Sections two and three focused on the frequency of the use of spirometry (intended as a measurement of airway flows and/or inference of respiratory volumes) in spontaneously ventilating and mechanically ventilated dogs. Section four asked if respondents considered spirometry “essential”. Section five asked what spirometry devices were used and about equipment maintenance/calibration. Section six focused on ventilation modes used and initial setup, including inspiratory pressure or tidal volume (VT) and adaptation to body condition score. Section seven gathered information on spirometry displays used, and, section eight, on  $C_{RS}$  and  $R_{RS}$ .

The questionnaire is available on request to the corresponding author.

## 2.8. Statistical Analysis

Time to answer the survey was not monitored. Every questionnaire was analysed. Completeness was checked for each question. Items were not weighted. Simple descriptive analysis was used; the percentage rate was calculated.

For VT analysis, respondents were categorized as either “anaesthetists” or “intensivists” according to their answer on which professional background (diploma, practice) they declared. Responses from people working in both specialties, blank responses, and “not applicable” were not included. Distribution of proposed VT values was checked using the d’Agostino normality test. Medians of VT were visually compared. To check whether intensivists and anaesthetists tended to use different VT, the Mann–Whitney *U*-test was performed (the middle of the range was arbitrarily taken as a single VT value for responses reporting a range rather than a single value). Varga and Delanay’s *A* was used to investigate the effect size. Successively, VT was defined as “low”, 8 mL kg<sup>-1</sup> or below, and “high”, 12 mL kg<sup>-1</sup> or above; if a range was centred around 8 or 12 mL kg<sup>-1</sup>, the lowest value was considered for the lower end of VT and the highest for the higher end so that “6 to 10 mL kg<sup>-1</sup>” would be classified as “low” and “10 to 15 mL kg<sup>-1</sup>” would be classified as “high”. The chi-square test (with Yates’ continuity correction) was used to test for differences in low versus high VT use between intensivists and anaesthetists. Analyses were performed using R V. 1.3.5 Mac OS (The R Foundation for Statistical Computing, <http://www.R-project.org>, Austria, accessed on 17 January 2022). A *p* value < 0.05 was considered significant.

## 3. Results

Section one, demographic information: There were 128 respondents (Table 1).

Sections two, three, and four, use of spirometry: In spontaneously ventilating dogs, spirometry was “never used” by 38/127 (30%) respondents [25/38 had no spirometer available; 13/38 had spirometers but were not using them] (no answer: 1/128). Spirometry was used by 27/127 (21%) of the respondents “in less than 25% of cases”, 11/127 (9%) “in 25 to 50% of cases”, 19/127 (15%) “in 50 to 75% of cases”, and 32/127 (25%) “in 75 to 100% of cases”.

**Table 1.** Demographics of the 128 respondents to a survey on the use of spirometry in small animal anaesthesia and critical care.

| Background                                  |  |              |                                     |
|---|--|--------------|-------------------------------------|
| Anaesthetists                               |  | Intensivists |                                     |
| Number                                      |  | Number       |                                     |
| 10  | Technicians/anaesthesia nurses   | 15           | ACVECC/ECVECC Diplomates            |
| 57  | ACVAA/ECVAA Diplomates   | 1            | ACVECC and ABVP Diplomate, and FCCM |
| 2   | Post ACVAA/ECVAA residency<br>(not Diplomates)   |              |                                     |
| 15  | ACVAA/ECVAA residents  | 1            | ACVECC/ECVECC resident              |
| 13  | GP with interest in anaesthesia  | 2            | GP with interest in ECC             |
| 1   | PhD student in anaesthesia   |              |                                     |
| Number                                      |  |              |                                     |
| 1   | ACVECC/ECVECC Diplomate, ACVAA/ECVAA resident  |              |                                     |
| 1   | ACVECC/ECVECC and ACVAA/ECVAA resident   |              |                                     |
| 4   | GP with interest in both anaesthesia and ECC   |              |                                     |
| 5   | Respondents being veterinarians with university teaching roles in anaesthesia and/or ECC but who did not match the abovementioned categories (one had a PhD) |              |                                     |
| Country                                     |  |              |                                     |
| United Kingdom (34/128)                     |  |              |                                     |
| United States of America (30/128)           |  |              |                                     |
| Australia (13/128)                          |  |              |                                     |
| Spain (12/128)                              |  |              |                                     |
| Switzerland (10/128)                        |  |              |                                     |
| Italy (9/128)                               |  |              |                                     |
| Austria (4/128)                             |  |              |                                     |
| France (4/128)                              |  |              |                                     |
| Canada (3/128)                              |  |              |                                     |
| Argentina (2/128)                           |  |              |                                     |
| Colombia (1/128)                            |  |              |                                     |
| Germany (1/128)                             |  |              |                                     |
| New-Zealand (1/128)                         |  |              |                                     |
| Norway (1/128)                              |  |              |                                     |
| South Africa (1/128)                        |  |              |                                     |
| South Korea (1/128)                         |  |              |                                     |
| Sweden (1/128)                              |  |              |                                     |
| Working environment                         |  |              |                                     |
| University Teaching Hospitals (69/128, 54%) |  |              |                                     |
| Referral centres (37/128, 29%)              |  |              |                                     |
| Private practice (17/128, 13%)              |  |              |                                     |
| Research institutions (4/128)               |  |              |                                     |
| Telemedicine (1/128).                       |  |              |                                     |

ABVP: American Board of Veterinary Practitioners; ACVAA: American College of Veterinary Anaesthesia and Analgesia; ACVECC: American College of Veterinary Emergency and Critical Care; ECC: Emergency and Critical care; ECVAA: European College of Veterinary Anaesthesia and Analgesia; ECVAA: European College of Veterinary Anaesthesia and Analgesia; FCCM: Fellow in the American College of Critical Care Medicine (ACCM); GP: General Practitioner.

In mechanically ventilated dogs, spirometry was “never used” by 24/127 (19%) respondents [18/24 had no spirometer available, 6/24 had access to spirometers] (no answer: 1/128). Spirometry was used by 15/127 (12%) of the respondents “in less than 25% of cases”, 10/127 (8%) “in 25 to 50% of cases”, 17/127 (13%) “in 50 to 75% of cases”, and 61/127 (48%) “in 75 to 100% of cases”.

All the 24/127 respondents that were not using spirometry in ventilated dogs also answered that they did not use spirometry in spontaneously ventilating dogs.

Spirometry was considered essential “in no case” by 11/128 respondents (9%), “in specific cases” by 55/128 respondents (43%), “in many cases” by 19/128 respondents (15%), and “in most cases” by 43/128 respondents (33%).

Section five, equipment and calibration: Most respondents (64/120, 53%) used the D-Lite/Pedi-Lite flow sensor (Pitot tube, Datex-Ohmeda/GE Healthcare). Wright respirometer, Spiromed (Dräger), or NICO (Respironics) were used by 18, 5, and 3 respondents, respectively. Other technologies, including heated wire anemometer, variable orifice flow sensor, ultrasonic flow sensor, and fixed orifice flow sensor (mostly non-specified), were used by 9, 7, 4, and 3 respondents, respectively. The ventilator display was used by 51/120 respondents (43%) to monitor spirometry. A variety of ventilators were used to monitor spirometry, but many respondents did not provide detail on their device(s) or on the technology behind the measurements. Four respondents used at least one additional technique to the above monitors/technologies but did not specify which. Six respondents did not know which technology they were using, of whom five considered spirometry necessary “in many” or “in most cases”. No answer: 8/128.

When asking about the verification of the calibration of spirometry devices, 65/119 (55%) declared that it was performed by the manufacturer when the device was serviced; 15/119 (13%) that it was regularly done by an anaesthesia team member; 14/119 declared that it was never checked (12%); and 25 (21%) reported they did not know. No answer: 9/128.

Section six, Ventilation settings: Pressure modes were used by 94/124 respondents (76%), volume modes by 104/124 (84%), and dual modes (e.g., pressure-regulated volume-controlled ventilation) by 49/124 (40%) (Multiple answers were possible; no answer: 4/128).

Regarding pressure modes, 10/128 respondents did not answer or did not use them. The initial pressure was set  $\leq 10$  cmH<sub>2</sub>O by 25/118 (21%) respondents (including 8 cmH<sub>2</sub>O by 11/118), at 10 cmH<sub>2</sub>O by 41/118 (35%), to a range centred around 10 cmH<sub>2</sub>O (pressure ranges: 9–11, 8–12, 5–15, 0–20) by 5/118 (4%), to a range  $\geq 10$  cmH<sub>2</sub>O by 47/118 (40%) (including 16/118 setting up 12 cmH<sub>2</sub>O and 6/118, 15 cmH<sub>2</sub>O). The pressure setting was adapted to size, breed, morphology, and BCS by 73/123 (59%) respondents (No answer: 5/128).

Regarding volume modes, 9/128 did not answer or did not use them. The initial VT was set  $\leq 10$  mL kg<sup>-1</sup> by 21/119 (18%) respondents, at 10 mL kg<sup>-1</sup> by 66/119 (55%), between 10 to 20 mL kg<sup>-1</sup> by 31/119 (26%) (including 10 using 12 mL kg<sup>-1</sup> and 5 using 15 mL kg<sup>-1</sup>), and between 8 to 12 mL kg<sup>-1</sup> by 1/119. The VT was adapted to size, breed, morphology, and body condition score (BCS) by 74/123 (58%) respondents (No answer: 5/128).


Median VT initially set by intensivist and anaesthetists was 9 mL kg<sup>-1</sup> and 10 mL kg<sup>-1</sup>, respectively ( $p < 0.01$ ; effect size was 0.78 [large]). A higher proportion of intensivists used “low” VT and a higher proportion of anaesthetists used “high” VT,  $\lambda^2(1, N = 40) = 15.09$ ,  $p < 0.01$ .

Section seven, spirometry displays: Pressure-volume loops were used by 89/117 respondents (76%), flow-volume loops by 51/117 (44%), pressure-time curves by 46/117 (39%), flow-time curves by 38/117 (32%), and volume-time curves by 20/117 (17%) (Multiple answers were possible; no answer: 11/128). A total of 18/117 respondents declared using no curve but only numbers (15%).

Section eight, C<sub>RS</sub> and R<sub>RS</sub>: Spirometry was “never used” specifically to monitor C<sub>RS</sub> and R<sub>RS</sub> by 26/125 (21%) respondents, used in “less than 25% of cases” by 34/125 (27%), used “in 25% to 50% of cases” by 10/125 (8%), used “in 50% to 75% of cases” by 17/125 (14%), and used “in 75% to 100% of cases” by 38/125 (30%) (no answer: 3/128).

Expected physiologic numerical values of C<sub>RS</sub> and R<sub>RS</sub> were extremely variable [80/128 (63%) and 106/128 (83%) respondents did not answer or stated they did not know].

A selection of results is summarized in Figure 1.



**Figure 1.** Summary of selected responses to an online survey on the use of spirometry in small animal anaesthesia and critical care by 128 respondents. (a) Number of respondents using spirometry in spontaneously ventilating dogs; (b) Number of respondents using spirometry in mechanically ventilated dogs; (c) Number of respondents considering spirometry essential; (d) Percentage of respondents using various spirometry displays.

#### 4. Discussion

Results of this survey illustrate that a variety of spirometry technologies was used in dogs, more frequently during mechanical ventilation than in spontaneous breathing. Most

respondents were not involved in or informed about the calibration of their equipment. Initial settings of mechanical ventilation varied widely. The shape of pressure-volume loops was the most commonly assessed display. Values of  $C_{RS}$  and  $R_{RS}$  remain unknown.

Over 3/4 of the respondents considered spirometry essential in “selected” or “most” cases. However, one-third of them was unaware of the calibration status of their equipment and over half relied exclusively on the service. The American Thoracic Society and the European Respiratory Society recommend daily checks of diagnostic spirometers (single discharge of a 3-L calibration syringe) and recalibration if the reading was out of the  $\pm 3\%$  range [4]. The same accuracy cannot be expected from spirometers used in anaesthesia (continuous monitoring, variation in gas composition). The manual of use for GE modules reports that the need for calibration is “infrequent” and that calibration should be performed only if the accuracy of volumes obtained was out of  $\pm 6\%$ . No evidence was found on the actual performance over time of spirometers used in anaesthesia. This should be investigated.

Respondents used the shape of the Pressure-Volume loop more than  $C_{RS}$  and  $R_{RS}$  absolute values. Pressure-volume loops were first described as sigmoid static curves [5]. Dynamic proximal (to the endotracheal tube) loops might be more useful than static curves to assess compromised airway patency, incorrect placements of the orotracheal tubes, leaks, or changes in  $C_{RS}$  and/or  $R_{RS}$  [6]. The resistance generated by the endotracheal tube can make their interpretation difficult [7]. Dynamic Pressure-Volume loops allow the monitoring of trends rather than precise measurements. The shape of such curves widely depends on the scale of the graph displayed, which can be altered or automatically adjusted by the monitors. Therefore, clinical interpretation of the respiratory mechanics could be incorrect. Values of  $C_{RS}$  and  $R_{RS}$  are independent of the scale of the curves displayed. In this survey, Pressure-Volume loops were used by one-third of the respondents in >75% of the cases. However, nearly 50% of all respondents did not use numerical values at all or used them in less than 25% of the cases. To date, equations only accounting for body weight have been reported for  $C_{RS}$  and  $R_{RS}$  values in healthy dogs [8,9]. Additional information and reference intervals are necessary to allow objective data interpretation. Additionally,  $C_{RS}$  and  $R_{RS}$  values can vary depending on the ventilation mode used [2].

Although nearly 60% of people adjusted their VT or inspiratory pressures in function of size, breed, morphology, and BCS, over half of the respondents set the initial VT at  $10 \text{ mL kg}^{-1}$  in volume-controlled ventilation. While (unreferenced) recommendation exists to start volume-controlled mechanical ventilation with VT below  $10 \text{ mL kg}^{-1}$  [10], evidence suggests that higher VT might be more appropriate in healthy dogs [11]. Interestingly, intensivists were more likely to use VT below  $8 \text{ mL kg}^{-1}$  and anaesthetists to use VT above  $12 \text{ mL kg}^{-1}$ . This may originate from differences in the animals' presentation in the different specialties (i.e., mostly healthy lungs in anesthesia vs. pathologic changes in ECC). It may also be that anesthetists and intensivists base their clinical decisions on different scientific literature [11,12]. Therefore, it might be interesting to differentiate the specialties in studies investigating respiratory mechanics.

This study has several limitations. Firstly, the number of respondents was low, and intensivists were under-represented. This could be due to the fact that the right channels were not used to contact the ECC community. This certainly limits the interpretation of our findings. Secondly, convenience sampling was used. Selection bias is likely: Respondents were probably interested in the topic to answer the survey. Thirdly, multiple responses from the same individual could not be excluded. Fourthly, some respondents stated they never used spirometry but gave detail on monitors and curves used. Possible explanations include either an error in entering the information or that these individuals work at centres where spirometers are available, but they do not use them clinically and interpreted the question asked in the survey in this regard. Fifthly, there were differences in the number of respondents not having access to a spirometer between spontaneous and ventilated dogs. This may reflect different contexts of spontaneous vs. mechanical ventilation and equipment availability.

## 5. Conclusions

Spirometry was used more frequently in mechanically ventilated dogs than during spontaneous ventilation by the respondents of this survey. The shape of pressure-volume loops was the most common display. A variety of technologies and ventilation modes were used. Future investigations should (1) assess the accuracy of spirometers commonly used in veterinary anaesthesia and (2) determine the reference intervals for  $C_{RS}$  and  $R_{RS}$  in healthy, anaesthetized dogs, as this survey confirmed the need for such information.

**Author Contributions:** Conceptualization, M.R., O.L. and M.M.; methodology, M.R., O.L. and M.M.; formal analysis, M.R. and O.L.; writing—original draft preparation, M.R.; writing—review and editing, M.R., O.L. and M.M.; supervision, O.L. and M.M. All authors have read and agreed to the published version of the manuscript.

**Funding:** This research received no external funding.

**Institutional Review Board Statement:** The study was conducted according to the guidelines of the Declaration of Helsinki. Ethical approval was waived for this study after review from the local ethical committee (Gesundheits-, Sozial- und Integrationsdirektion Kantonale Ethikkommission für die Forschung, Kanton Bern) because this project was not subject to any authorization according to national law (Humanforschungsgesetz, Art. 2, Abs 1).

**Informed Consent Statement:** Informed consent was obtained from all subjects involved in the study. The introduction page informed the participants on the objectives, number of questions, and expected time needed to answer the survey. It also reported: “The data collected will remain confidential. No information to identify you or your practice will be collected. (...) You are undertaking this survey on a voluntary basis. You can decide at any time that you no longer want to participate. (...) By clicking “NEXT” you certify that you have read and understood the purpose of this study, your rights as an informed participant and that your answers will be used for analysis and publication.”

**Data Availability Statement:** The data is available on reasonable request to the corresponding author.

**Acknowledgments:** The authors would like to thank the AVA (Association of Veterinary Anaesthetists), the SEAAV (Sociedad Española de Anestesia y Analgesia Veterinaria), the Anaesthesia and Emergency and Critical care chapters of the ANZCVS (Australia and New Zealand College of Veterinary Scientists), the EVECCS (European Veterinary Emergency and Critical Care Society), the ECVA (European College of Veterinary Anaesthesia and Analgesia), the ACVECC (American College of Emergency and Critical Care), and the administrators of the acva-list for their help with the survey advertisement.

**Conflicts of Interest:** The authors declare no conflict of interest.



## Appendix A. A Selection of Survey Announcements

**Association of Veterinary Anaesthetists**  
3 février 2018 · 🌐

Survey into the use of spirometry in Veterinary Anaesthesia

**Evaluation of the current use of spirometry in Veterinary Anaesthesia and Intensive care**

Dear Colleague,

Thank you for your interest in this study.

Thorough knowledge and understanding of respiratory mechanics is essential to guide clinical decisions in mechanically ventilated animals. Our aim is to characterize global compliance as well as resistance of the respiratory system in anaesthetized dogs.

**Purpose:** As a first step, we wish to document the current use of spirometry within the Veterinary Anaesthesia and Intensive care community. Our intention is to survey opinions and expectations of veterinary practitioners, anaesthetists and intensivists on spirometry with particular emphasis on respiratory compliance and resistance.

**Procedure:** This questionnaire is expected to take you between 5 and 10 minutes to complete. Your opinion is relevant, there are no "right" or "wrong" answers. Following the questionnaire, you will be asked if your centre wishes to collaborate in a multicentre data collection at a later stage.

**Confidentiality:** The data collected will remain confidential. No information to identify you or your practice will be collected. Should you be willing to be contacted for the inclusion in the multicentre study, your email address will be recorded but will not be associated with your survey responses.

DOCS.GOOGLE.COM


**Evaluation of the current use of spirometry in Veterinary Anaesthesia and Intensive care**

[Boostez la publication](#)

👍 9 1 partage

👍 J'aime    💬 Commenter    ➦ Partager    AVA ▼

**Figure A1.** Example of survey announcement on the Facebook page of the Association of Veterinary Anaesthetists (AVA).



**SEAAV**  
2 février 2018 · 🌐

Nuestro amigo [Mathieu Raillard](#), de la Universidad de Berna nos pide colaboración en un estudio sobre espirometría. Se trata de rellenar una encuesta cortita online.  
Anímate y hazla. No se tardan más de 5 minutos.

Dear all,  
We hope you are not too tired of completing online surveys at the moment. We would be very grateful if, ones again, you could take the time to help us getting valuable data. We would like to document the current use of spirometry within the Veterinary Anaesthesia and Intensive care community. This questionnaire is expected to take 5 - 10 minutes to complete.  
If you wish to participate please click on the web link below or copy it into your internet browser:  
<https://docs.google.com/.../1FAIpQLSe-tt48Oyi5D28xe.../viewform...>  
Following the questionnaire, you will be asked if your centre wishes to collaborate in a multicentre data collection aimed to characterize global compliance and resistance of the respiratory system in anaesthetized dogs at a later stage.  
Feel free to share this link with anyone not on the list.  
Thank you,  
Mathieu Raillard, Olivier Levionnois & Martina Mosing

**Evaluation of the current use of spirometry in Veterinary Anaesthesia and Intensive care**

Dear Colleague,  
Thank you for your interest in this study.

**Thorough knowledge and understanding of respiratory mechanics is essential to guide clinical decisions in mechanically ventilated animals. Our aim is to characterize global compliance as well as resistance of the respiratory system in anaesthetized dogs.**

**Purpose:** As a first step, we wish to document the current use of spirometry within the Veterinary Anaesthesia and Intensive care community. Our intention is to survey opinions and expectations of veterinary practitioners, anaesthetists and intensivists on spirometry with particular emphasis on respiratory compliance and resistance.

**Procedure:** This questionnaire is expected to take you between 5 and 10 minutes to complete. Your opinion is relevant, there are no "right" or "wrong" answers. Following the questionnaire, you will be asked if your centre wishes to collaborate in a multicentre data collection at a later stage.


**Confidentiality:** The data collected will remain confidential. No information to identify you or your practice will be collected. Should you be willing to be contacted for the inclusion in the multicentre study, your email address will be recorded but will not be associated with your survey responses.

DOCS.GOOGLE.COM  
**Evaluation of the current use of spirometry in Veterinary Anaesthesia and Intensive care**

👍 29 8 partages

👍 J'aime    💬 Commenter    ➦ Partager    🌐

**Figure A2.** Example of survey announcement on the Facebook page of the Sociedad Española de Anestesia y Analgesia Veterinaria (SEAAV).



**Anzcvs Veterinary Anaesthesia & Analgesia Chapter**  
7 février 2018 · 🌐

**YOUR ASSISTANCE IS NEEDED!**

Please see below a request from Dr Raillard and colleagues for help with a survey into the use of spirometry in veterinary anaesthesia and intensive care:

Dear all,

We hope you are not too tired of completing online surveys at the moment. We would be very grateful if, once again, you could take the time to help us getting valuable data. We would like to document the current use of spirometry within the Veterinary Anaesthesia and Intensive care community. This questionnaire is expected to take 5 - 10 minutes to complete.

If you wish to participate please click on the web link below or copy it into your internet browser:

<https://docs.google.com/.../1FAIpQLSe-tt48Oyi5D28xe.../viewform...>

Following the questionnaire, you will be asked if your centre wishes to collaborate in a multicentre data collection aimed to characterize global compliance and resistance of the respiratory system in anaesthetized dogs at a later stage.

Thank you,

Mathieu Raillard, DipECVAA, MRCVS  
naesthetist at Vetsuisse-Faculty University Bern, Switzerland

Olivier Levionnois, Dr. Med. Vet, PhD, DipECVAA  
Senior Lecturer at Vetsuisse-Faculty University Bern, Switzerland

Martina Mosing, Dr. Med. Vet, PhD, DipECVAA  
Senior Lecturer at School of Veterinary and Life Sciences, Murdoch University, Perth, Western Australia

**Evaluation of the current use of spirometry in Veterinary Anaesthesia and Intensive care**

Dear Colleague,

Thank you for your interest in this study.

Thorough knowledge and understanding of respiratory mechanics is essential to guide clinical decisions in mechanically ventilated animals. Our aim is to characterize global compliance as well as resistance of the respiratory system in anaesthetized dogs.

**Purpose:** As a first step, we wish to document the current use of spirometry within the Veterinary Anaesthesia and intensive care community. Our intention is to survey opinions and expectations of veterinary practitioners, anaesthetists and intensivists on spirometry with particular emphasis on respiratory compliance and resistance.

**Procedure:** This questionnaire is expected to take you between 5 and 10 minutes to complete. Your opinion is relevant, there are no "right" or "wrong" answers. Following the questionnaire, you will be asked if your centre wishes to collaborate in a multicentre data collection at a later stage.

**Confidentiality:** The data collected will remain confidential. No information to identify you or your practice will be collected. Should you be willing to be contacted for the inclusion in the multicentre study, your email address will be recorded but will not be associated with your survey responses.

**Withdrawal of participation:** You are undertaking this survey on a voluntary basis. You can decide at any time that you no longer want to participate.

DOCS.GOOGLE.COM

**Evaluation of the current use of spirometry in Veterinary Anaesthesia and Intensive care**

👍 Martina Mosing et 1 autre personne

1 partage

**Figure A3.** Example of survey announcement on the Facebook page of the Anzcvs Veterinary Anaesthesia & Analgesia Chapter.

**Veterinary Anaesthesia Tips**  
7 février 2018 · 🌐

**Anzcvp Veterinary Anaesthesia & Analgesia Chapter**  
7 février 2018 · 🌐 J'aime la Page

**YOUR ASSISTANCE IS NEEDED!**

Please see below a request from Dr Raillard and colleagues for help with a survey into the use of spirometry in veterinary anaesthesia and intensive care:

Dear all,

We hope you are not too tired of completing online surveys at the moment. We would be very grateful if, once again, you could take the time to help us getting valuable data. We would like to document the current use of spirometry within the Veterinary Anaesthesia and Intensive care community. This questionnaire is expected to take 5 - 10 minutes to complete.

If you wish to participate please click on the web link below or copy it into your internet browser:

<https://docs.google.com/.../1FAIpQLSe-tt48Oyi5D28xe.../viewform...>

Following the questionnaire, you will be asked if your centre wishes to collaborate in a multicentre data collection aimed to characterize global compliance and resistance of the respiratory system in anaesthetized dogs at a later stage.

Thank you,

Mathieu Raillard, DipECVAA, MRCVS  
naesthetist at Vetsuisse-Faculty University Bern, Switzerland

Olivier Levionnois, Dr. Med. Vet, PhD, DipECVAA  
Senior Lecturer at Vetsuisse-Faculty University Bern, Switzerland

Martina Mosing, Dr. Med. Vet, PhD, DipECVAA  
Senior Lecturer at School of Veterinary and Life Sciences, Murdoch University, Perth, Western Australia

**Evaluation of the current use of spirometry in Veterinary Anaesthesia and Intensive care**

Dear Colleagues,

Thank you for your interest in this study.

Thorough knowledge and understanding of respiratory mechanics is essential to guide clinical decisions in mechanically ventilated animals. Our aim is to characterize global compliance as well as resistance of the respiratory system in anaesthetized dogs.

Purpose: As a first step, we wish to document the current use of spirometry within the Veterinary Anaesthesia and Intensive care community. Our attention is to survey opinions and expectations of veterinary practitioners, anaesthetists and intensivists on spirometry with particular emphasis on respiratory compliance and resistance.

Procedure: This questionnaire is expected to take you between 5 and 10 minutes to complete. Your opinion is relevant, there are no "right" or "wrong" answers. Following the questionnaire, you will be asked if your centre wishes to collaborate in a multicentre data collection at a later stage.

Confidentiality: The data collected will remain confidential. No information to identify you or your practice will be collected. Should you be willing to be contacted for the inclusion in the multicentre study, your email address will be recorded but will not be associated with your survey responses.


Withdrawal of participation: You are undertaking this survey on a voluntary basis. You can decide at any time that you no longer want to participate.


DOCS GOOGI E COM  
**Evaluation of the current use of spirometry in Veterinary Anaesthesia and Intensive care**

👍 3

J'aime Commenter Partager

**Figure A4.** Example of survey announcement on the Facebook page of the Veterinary Anaesthesia Tips.



 **European Veterinary Emergency and Critical Care Society** ...

13 février 2018 · 🌐

Dear all,  
Please help our friends from AVA to get usefull information of spirometry role in anaesthesia and intensive care pratice.  
If you wish to participate please click on the web link below or copy it into your internet browser:  
<https://docs.google.com/.../1FAIpQLSe-tt48Oyi5D28xe.../viewform...>  
Thnak you all

**Evaluation of the current use of spirometry in Veterinary Anaesthesia and Intensive care**

Dear Colleague,

Thank you for your interest in this study.

Thorough knowledge and understanding of respiratory mechanics is essential to guide clinical decisions in mechanically ventilated animals. Our aim is to characterize global compliance as well as resistance of the respiratory system in anaesthetized dogs.

**Purpose:** As a first step, we wish to document the current use of spirometry within the Veterinary Anaesthesia and Intensive care community. Our intention is to survey opinions and expectations of veterinary practitioners, anaesthetists and intensivists on spirometry with particular emphasis on respiratory compliance and resistance.


**Procedure:** This questionnaire is expected to take you between 5 and 10 minutes to complete. Your opinion is relevant, there are no "right" or "wrong" answers. Following the questionnaire, you will be asked if your centre wishes to collaborate in a multicentre data collection at a later stage.





**Confidentiality:** The data collected will remain confidential. No information to identify you or your practice will be collected. Should you be willing to be contacted for the inclusion in the multicentre study, your email address will be recorded but will not be associated with your survey responses.

**Withdrawal of participation:** You are undertaking this survey on a voluntary basis. You can decide at any time that you no longer want to participate.


DOCS.GOOGLE.COM

**Evaluation of the current use of spirometry in Veterinary Anaesthesia and Intensive care**

 4 2 partages


 J'aime  Commenter  Partager 

**Figure A5.** Example of survey announcement on the Facebook page of the European Veterinary Emergency and Critical Care Society.



The screenshot shows a web browser window displaying the AVA website. The main content is a survey announcement titled "SURVEY ON THE CURRENT USE OF SPIROMETRY IN VETERINARY ANAESTHESIA AND INTENSIVE CARE". The announcement is published by Alex Chebroux on February 4, 2018, at 8:46 pm. The text invites the community to participate in a questionnaire about spirometry use, with a link to a Google Forms survey. It also mentions that participants will be asked if their centres wish to collaborate in a multicentre data collection project. The announcement is signed by Mathieu Raillard, Olivier Levionnois, and Martina Mosing. There are social media sharing buttons for Twitter, Facebook, and Google+. The page is categorized under "Latest News, Surveys". On the right side, there is a "RELATED STORIES" section with three items: "Online survey about your anaesthesia residency programme (ACVAA or ECVAA)", "Kate Borer-Weir Memorial Fund", and "Survey on anaesthetic protocols for small animal caesarean sections". The footer contains copyright information for the Association of Veterinary Anaesthetists (2020), social media icons, and a list of links: Constitution, Byelaws, Privacy Notice, Terms, and Contact. The website is powered by Wahoo Group.

**Figure A6.** Example of survey announcement on the Association of Veterinary Anaesthetists (AVA) website.



**Figure A7.** Example of survey announcement on the ACVA-L.

## References

1. Calice, I.; Moens, Y. Modern Spirometry Supports Anesthetic Management in Small Animal Clinical Practice: A Case Series. *J. Am. Anim. Hosp. Assoc.* **2016**, *52*, 305–311. [[CrossRef](#)] [[PubMed](#)]
2. Fantoni, D.T.; Ida, K.K.; Lopes, T.F.; Otsuki, D.A.; Auler, J.O., Jr.; Ambrosio, A.M. A comparison of the cardiopulmonary effects of pressure controlled ventilation and volume controlled ventilation in healthy anesthetized dogs. *J. Vet. Emerg. Crit. Care* **2016**, *26*, 524–530. [[CrossRef](#)] [[PubMed](#)]

3. Eysenbach, G. Improving the quality of Web surveys: The Checklist for Reporting Results of Internet E-Surveys (CHERRIES). *J. Med. Internet Res.* **2004**, *6*, e34. [[CrossRef](#)] [[PubMed](#)]
4. Miller, M.R.; Hankinson, J.; Brusasco, V.; Burgos, F.; Casaburi, R.; Coates, A.; Crapo, R.; Enright, P.; van der Grinten, C.P.; Gustafsson, P.; et al. Standardisation of spirometry. *Eur. Respir. J.* **2005**, *26*, 319–338. [[CrossRef](#)] [[PubMed](#)]
5. Rahn, H.; Otis, A.B.; Chadwick, L.E.; Fenn, W.O. The pressure-volume diagram of the thorax and lung. *Am. J. Physiol.* **1946**, *146*, 161–178. [[CrossRef](#)] [[PubMed](#)]
6. Bardoczky, G.I.; Engelman, E.; d'Hollander, A. Continuous spirometry: An aid to monitoring ventilation during operation. *Br. J. Anaesth.* **1993**, *71*, 747–751. [[CrossRef](#)] [[PubMed](#)]
7. Stenqvist, O. Practical assessment of respiratory mechanics. *Br. J. Anaesth.* **2003**, *91*, 92–105. [[CrossRef](#)] [[PubMed](#)]
8. Bradbrook, C.A.; Clark, L.; Dugdale, A.H.; Burford, J.; Mosing, M. Measurement of respiratory system compliance and respiratory system resistance in healthy dogs undergoing general anaesthesia for elective orthopaedic procedures. *Vet. Anaesth. Analg.* **2013**, *40*, 382–389. [[CrossRef](#)] [[PubMed](#)]
9. Asorey, I.; Pellegrini, L.; Canfran, S.; Ortiz-Diez, G.; Aguado, D. Factors affecting respiratory system compliance in anaesthetised mechanically ventilated healthy dogs: A retrospective study. *J. Small Anim. Pract.* **2020**, *61*, 617–623. [[CrossRef](#)] [[PubMed](#)]
10. Hopper, K.; Powell, L.L. Basics of mechanical ventilation for dogs and cats. *Vet. Clin. N. Am. Small Anim. Pract.* **2013**, *43*, 955–969. [[CrossRef](#)] [[PubMed](#)]
11. Bumbacher, S.; Schramel, J.P.; Mosing, M. Evaluation of three tidal volumes (10, 12 and 15 mL kg<sup>-1</sup>) in dogs for controlled mechanical ventilation assessed by volumetric capnography: A randomized clinical trial. *Vet. Anaesth. Analg.* **2017**, *44*, 775–784. [[CrossRef](#)] [[PubMed](#)]
12. Oura, T.; Rozanski, E.A.; Buckley, G.; Bedenice, D. Low tidal volume ventilation in healthy dogs. *J. Vet. Emerg. Crit. Care* **2012**, *22*, 368–371. [[CrossRef](#)] [[PubMed](#)]