

EDITORIAL

Open Access



Collection on total-body PET

Kuangyu Shi^{1*} and Charalampos Tsoumpas²

*Correspondence:
kuangyu.shi@unibe.ch

¹ Department of Nuclear
Medicine, Bern University
Hospital, University of Bern, Bern,
Switzerland

² Department of Nuclear
Medicine and Molecular
Imaging, University
of Groningen, University Medical
Center Groningen, Groningen,
The Netherlands

With the prompt appearance of extended axial field of view positron emission tomographs (LAFOV PET), an extensive range of prospects become possible in molecular diagnostics. An advantage using LAFOV PET is the increase of sensitivity, which can be more than 10 times higher for 1-m-long PET [1] and 40 times for a 2-m-long PET scanner [2], yet it is essential to state that these figures depend on the patient, organs of study and tracer distribution.

On the one hand, the axial extension of the scanners and their increased sensitivity offer a mixture of benefits:

- Ultra-fast scanning [3]
- Simultaneous imaging of more organs [4]
- Scanning kinetics with wider temporal range (e.g. delayed imaging or very short frames) [5]
- Enhanced image quality [6]
- Scanning with appreciably lower injected radioactivity [7]
- Enhanced support of systems biology/medicine [8]
- Enhanced support of dosimetry and theranostics [9]

On the other hand, LAFOV PET scanners also brought new methodological challenges [10]. The tremendous numbers of detectors strongly increased the complexity and data volume. It becomes more challenging to deal with large heterogeneity inside the scanner, for example, due to the oblique lines of response.

Several LAFOV PET scanners across the world have been already installed and been in use creating a wide range of exciting research [11].

EJNMMI also offers a collection on total body PET [11] which is available at <https://link.springer.com/journal/259/collections>.

We look forward to receiving submissions of work by researchers from all around the world.

Published online: 10 July 2023



© The Author(s) 2023. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

References

1. van Sluis J, van Snick JH, Brouwers AH, Noordzij W, Dierckx RAJO, Borra RJH, Slart RHJA, Lammertsma AA, Glaudemans AWJM, Boellaard R, Tsoumpas C. EARL compliance and imaging optimisation on the Biograph Vision Quadra PET/CT using phantom and clinical data. *Eur J Nucl Med Mol Imaging*. 2022;49(13):4652–60. <https://doi.org/10.1007/s00259-022-05919-1>.
2. Spencer BA, Berg E, Schmall JP, Omidvari N, Leung EK, Abdelhafez YG, Tang S, Deng Z, Dong Y, Lv Y, Bao J, Liu W, Li H, Jones T, Badawi RD, Cherry SR. Performance evaluation of the uEXPLORER total-body PET/CT scanner based on NEMA NU 2–2018 with additional tests to characterize PET scanners with a long axial field of view. *J Nucl Med*. 2021;62(6):861–70. <https://doi.org/10.2967/jnumed.120.250597>.
3. Zhang Y, Hu P, He Y, Yu H, Tan H, Liu G, Gu J, Shi H. Ultrafast 30-s total-body PET/CT scan: a preliminary study. *Eur J Nucl Med Mol Imaging*. 2022;49(8):2504–13. <https://doi.org/10.1007/s00259-022-05838-1>.
4. Surti S, Pantel AR, Karp JS. Total body PET: why, how, what for? *IEEE Trans Radiat Plasma Med Sci*. 2020;4(3):283–92. <https://doi.org/10.1109/trpms.2020.2985403>.
5. Zhang X, Cherry SR, Xie Z, Shi H, Badawi RD, Qi J. Subsecond total-body imaging using ultrasensitive positron emission tomography. *Proc Natl Acad Sci U S A*. 2020;117(5):2265–7. <https://doi.org/10.1073/pnas.1917379117>.
6. Prenosil GA, Hentschel M, Weitzel T, Sari H, Shi K, Afshar-Oromieh A, Rominger A. EARL compliance measurements on the biograph vision Quadra PET/CT system with a long axial field of view. *EJNMMI Phys*. 2022;9(1):26. <https://doi.org/10.1186/s40658-022-00455-1>.
7. Djekidel M, AlSadi R, Akl MA, Vandenberghe S, Bouhali O. Total-body pediatric PET is ready for prime time. *Eur J Nucl Med Mol Imaging*. 2022;49(11):3624–6. <https://doi.org/10.1007/s00259-022-05873-y>.
8. Abdelhafez Y, Raychaudhuri SP, Mazza D, Sarkar S, Hunt HL, McBride K, Nguyen M, Caudle DT, Spencer BA, Omidvari N, Bang H, Cherry SR, Nardo L, Badawi RD, Chaudhari AJ. Total-Body ^{18}F -FDG PET/CT in autoimmune inflammatory arthritis at ultra-low dose: initial observations. *J Nucl Med*. 2022;63(10):1579–85. <https://doi.org/10.2967/jnumed.121.263774>.
9. Yu H, Lv J, Hu P, Chen S, Shi H. Reduction of radiation accumulation in salivary glands through oral vitamin C during ^{68}Ga -PSMA-11 total-body dynamic PET/CT imaging. *Nucl Med Commun*. 2022;43(2):166–71. <https://doi.org/10.1097/MNM.0000000000001506>.
10. Nadig V, Herrmann K, Mottaghy FM, Schulz V. Hybrid total-body pet scanners-current status and future perspectives. *Eur J Nucl Med Mol Imaging*. 2022;49(2):445–59. <https://doi.org/10.1007/s00259-021-05536-4>.
11. Slart RHJA, Rominger A. Collection issue on clinical total-body PET studies. *Eur J Nucl Med Mol Imaging*. 2023;50:1556–7.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Submit your manuscript to a SpringerOpen[®] journal and benefit from:

- Convenient online submission
- Rigorous peer review
- Open access: articles freely available online
- High visibility within the field
- Retaining the copyright to your article

Submit your next manuscript at ► [springeropen.com](https://www.springeropen.com)