

# Harvesting energy from the heart wall motion – Device weight considerations

A. Zurbuchen<sup>1</sup>, A. Haeberlin<sup>1,2</sup>, J. Schaerer<sup>1</sup>, A. Pfenniger<sup>1</sup>, R. Vogel<sup>1,3</sup>

<sup>1</sup> ARTORG Cardiovascular Engineering, University of Bern, Bern, Switzerland

<sup>2</sup> Department of Cardiology, Bern University Hospital and University of Bern, Bern, Switzerland

<sup>3</sup> Department of Cardiology, Bürgerspital Solothurn, Solothurn, Switzerland

## Introduction

Primary batteries are used as energy source in active medical implants. However, the limited lifetime of batteries requires repeated replacement procedures associated with risk-bearing surgical interventions. To provide a continuous intracorporeal energy supply, we developed an energy harvesting device converting cardiac motion into electrical energy. The device – a modified autoquarz wrist watch – was fixated directly on the epicardial side of the ventricular wall. The impact of this additional mass on the heart was investigated to optimise weight and energy conversion efficiency.

## Methods

A sternotomy was performed on a 60 kg domestic pig during an in vivo study. The energy harvesting device and a 9-axis inertial sensor were subsequently sutured onto the epicardium of the anteroapical segment of the left ventricle. The sensor was embedded into a plastic housing (4.5 g) carrying different additional loads from 0 to 20 g. A series of inertial measurements were acquired with a total load of 4.5 up to 24.5 g. These measurements were used as input for a mathematical model predicting the energy output of the harvesting device.

## Results

The harvesting device generated a mean output power of 52  $\mu$ W. Acceleration measurements show a change in ventricular wall motion as a result of different load conditions: the physical work of the heart increased from 6 mJ at 4.5 g to 45 mJ at 24.5 g. This additional effort leads to higher forces acting on the harvesting device and ultimately results in an increased energy conversion. Simulations estimate a drop in harvested energy of more than 60% for a weight reduction from 24.5 g to 4.5 g.

## Conclusion

Excessive overloading of the heart will interfere with cardiac function. However, we were able to load the heart with 24.5 g without observing any acute cardiac dysfunction. Furthermore, a heavier harvesting device favours the energy conversion.