

# Back to 1D Detectors: Benefits for High-Pressure Powder X-ray Diffraction

Authors

**Martin Fisch<sup>ab</sup>, Arianna Lanza<sup>ab</sup>, Piero Macchi<sup>a</sup> and Nicola Casati<sup>b\*</sup>**

<sup>a</sup>Department of Chemistry and Biochemistry, University of Bern, Freiestrasse 3, 3012 Bern, Switzerland

<sup>b</sup>Swiss Light Source, Paul Scherrer Institute, 5232 Villigen, Switzerland

Correspondence email: [nicola.casati@psi.ch](mailto:nicola.casati@psi.ch)

**Synopsis** A simple method to obtain high-resolution high-pressure powder X-ray diffraction data with accurate intensities from a 1D detector is presented and successfully applied to solving the structure of an L-Serine polymorph.

**Abstract** High-pressure powder X-ray diffraction is a fundamental technique for investigating structural responses to externally applied force. Synchrotron sources and 2D detectors are required. In contrast to this conventional setup, high-resolution beamlines equipped with 1D detectors could offer much better resolved peaks but cannot deliver accurate structure factors because they only sample a small portion of the Debye rings, which are usually inhomogeneous and spotty due to the small amount of sample. In this study, a simple method to overcome this problem is presented and successfully applied to solving the structure of an L-Serine polymorph from powder data. A comparison of the obtained high-resolution high-pressure data to conventional one shows that this technique, providing up to ten times better angular resolution, can be of advantage for indexing, lattice parameter refinement and even for structure refinement and solution in special cases.

**Keywords:** 1D Detector; High-resolution high-pressure powder X-ray diffraction; Accurate intensities.

## 1. Introduction

Increasing pressure on solid state materials can provide a significant amount of energy to such systems, which is capable of altering nuclear positions, generating new phases and eventually triggering chemical reactions. Investigating materials' behavior as a function of pressure is therefore essential to improve knowledge of structure – property relations and provides basis for a deeper understanding of many condensed matter phenomena.

With the invention of the diamond anvil cell (DAC, Jamieson *et al.*, 1959), high-pressure (HP) studies have become widely accessible. Whereas DACs can be used in other analytical methods such as optical microscopy and spectroscopy, diffraction is the elective technique to obtain complete information on the crystal structure at a given pressure and temperature.

The high information content of single-crystal diffraction data makes it the method of choice for solving and refining structures at HP and such experiments can nowadays be routinely performed with laboratory X-ray diffractometers.

Often, however, HP experiments cannot be performed on single crystalline samples, as they may not be available for a species or may experience severe fragmentation during first-order pressure-driven phase transitions. Moreover, powder samples may be preferred when the phase evolution and transition points with increasing pressure are different for single crystals or powders, or in time resolved investigations as experiments are faster (Boldyreva, 2007; Lanza *et al.*, 2014; Fisch *et al.*, 2015; Shekar & Rajan, 2001; Perillat, 2008; Evans *et al.*, 2007; Velisavljevic *et al.*, 2014).

HP powder X-ray diffraction (P-XRD) experiments are, however, challenging and prone to many difficulties. Among the technical problems are e.g. absorption from the DAC, Compton scattering (leading to high background), very intense Bragg peaks from the diamonds, scattering from the metal gasket and the restricted reciprocal space access due to shading from the DAC body. Moreover, at variance from traditional P-XRD experiments, random particle orientation, mandatory for extraction of reliable structure factors, is very difficult to achieve because of the very small amount of sample in a DAC (usually  $\leq 200 \mu\text{m}$  diameter and  $\leq 80 \mu\text{m}$  depth = ca.  $0.003 \text{ mm}^3$ ). In many cases, preferential orientation of the particles due to uniaxial stresses in the DAC (Tschauner *et al.*, 2005), further affects the intrinsically weak and spotty diffraction. For data of acceptable quality, synchrotron radiation sources are usually required.

Even though the number of crystallites in Bragg condition is increased when the DAC is oscillated a few degrees  $\pm \omega$  (Smith & Desgreniers, 2008), extraction of accurate intensities for structure solution or refinement from powder data still requires sampling of a large fraction of the Debye rings. Point or 1D detectors are not suitable for such applications, and a typical setup involves a sample stage coupled with 2D imaging plates, CCD/CMOS or photon counting pixel detectors with a high dynamic range to avoid saturation by high intensity reflections from the sample and especially the diamonds.

The frames are subsequently integrated and processed to a one dimensional diffraction pattern with software such as the widely used Fit2D (Hammersley *et al.*, 1996), GSAS-II (Toby & Von Dreele, 2013), XRD2DSCAN (Rodriguez-Navarro, 2006), Datasqueeze (Heiney, 2005) and various synchrotron in-house developments as e.g. Dioptas (Prescher & Prakapenka, 2015), Nika (Ilavsky, 2012) and DPDAK (Benecke *et al.*, 2014), many of which are based on the Python library pyFAI (Ashiotis *et al.*, 2015).

Two dimensional detectors are usually not exceeding few megapixels, for practical and cost reasons. In order to collect the whole circumference of Debye rings up to small d-values, they have to be placed relatively close to the sample. Consequently, the spatial resolution of adjacent rings is reduced and peaks tend to overlap.

In summary, there is a constant tradeoff between access to reciprocal space, angular resolution and hence the accuracy of peak positions and extracted structure factors in HP P-XRD experiments.

In this study we discuss the use of a 1D microstrip detector for this type of experiments, showing the advantages in resolution and describing a strategy for obtaining reliable intensities with such a setup. By comparing the obtained data with those measured with a 2D detector, we show the potential of this method for applications which require extremely high-resolution powder diffraction data at HP. This is showcased investigating L-Serine. An unsolved HP polymorph (L-Ser IV) was observed only in powder samples (Fisch *et al.*, 2015) and was always coexisting with L-Ser I (ambient form) and L-Ser II (an already known HP form). The large cell of L-Ser IV results in severe peak overlap with the other phases. Indexing, space-group assignment and extraction of structure factors for subsequent structure solution therefore profit from powder data with exceptionally high angular resolution.

## 2. Experimental

### 2.1. Sample and DAC preparation

Certified L-Serine (Sigma-Aldrich 54763, Lot# BCBH42640V) was ground in an agate mortar and further milled between finely ground glass plates to minimal particle size. Two samples (Table 1) were prepared in DACs with 0.5 mm and 0.6 mm culets. Steel discs with a hole of 0.25 mm, preindented to ca. 0.09 mm, served as gaskets and a 4:1 MeOH:EtOH solution as pressure medium. Pressure was raised with steps  $< 0.6(1)$  GPa each ca. 30 min as measured from ruby fluorescence (Syassen, 2008). Final pressure was determined from L-Serine lattice parameters (Fisch *et al.*, 2015). Furthermore, a DAC with 0.3 mm culets was loaded with NIST SRM 660a LaB<sub>6</sub> but not compressed (Table 1).

### 2.2. Beamline setups

Synchrotron powder XRD experiments were performed at the Swiss Light Source, Paul Scherrer Institute.

The X04SA Materials Science (MS) beamline's powder diffractometer (Willmott *et al.*, 2013) is equipped with a PSI Mythen II 1D single-photon counting multistrip detector (Bergamaschi *et al.*, 2010), mounted 0.76 m from the sample. It has a microstrip size of 50 (H) x 800 (W)  $\mu\text{m}^2$ , resulting in an intrinsic resolution smaller than  $0.0038^\circ 2\theta$ . The curved detector covers a  $2\theta$  range of  $120^\circ$  and was positioned to collect data from negative to positive diffraction angles. Radiation was monochromated to 16 keV and the primary beam focused and shaped to a square of 100 by 100  $\mu\text{m}^2$ . Wavelength and zero-offsets were fine calibrated to lattice parameters of NIST SRM 660a  $\text{LaB}_6$ . A Huber Eulerian Cradle 511 (Huber Diffractionstechnik GmbH & Co. KG, Germany) was mounted onto the  $\omega$  circle, allowing for  $\chi$  rotation of the DAC coaxial with the beam. After centering the cradle to the beam, the gasket hole was aligned to the center of the goniometer with a motorized xy stage on the cradle's  $\phi$  circle. An overview of the setup is shown in Fig. 1.

2D diffraction data were collected at the X06DA Macromolecular Crystallography (PXIII) beamline (Waltersperger *et al.*, 2015), equipped with a Dectris Pilatus 2M area detector (Broennimann *et al.*, 2006). The detector and pixel size are 254 (W) x 289 (H)  $\text{mm}^2$  and 172 x 172  $\mu\text{m}^2$ , respectively and the beam (16 keV) was focused and shaped to 90 (W) x 50 (H)  $\mu\text{m}^2$ . With the set detector distances of 0.2 m and 0.4 m, the angular coverage for complete Debye rings is  $32.4^\circ 2\theta$  and  $17.6^\circ 2\theta$ , respectively. The best resolution for fully covered rings is calculated as  $0.035^\circ 2\theta$  at 0.2 m and  $0.022^\circ 2\theta$  at 0.4 m (Hinrichsen *et al.*, 2008).

### 2.3. Data collection

Sample  $\text{LaB}_6$ -1D was measured at one fixed  $\chi$  position with the 1D detector and oscillated  $\pm 3^\circ \omega$  during data collection. The same DAC ( $\text{LaB}_6$ -2D) was measured with the 2D detector, with 0.2 m and with 0.4 m detector distances and moved from  $0^\circ \omega$  to  $4^\circ \omega$  while acquiring.

All DACs loaded with L-Serine (Table 1) were measured with the 1D detector. In addition, sample SerB-2D was measured with the 2D detector (0.2 m distance) and the same  $\omega$  rotation as the  $\text{LaB}_6$  DAC. The smaller detector distance ensures complete sampling of the Debye rings above  $20^\circ 2\theta$ . A quadrant of the 2D data is shown in Fig. 2.

On sample SerA- $\chi$ , 144 diffraction patterns were collected with the 1D detector each  $2.5^\circ \chi$  while oscillating  $\pm 6^\circ \omega$  during data acquisition. A quadrant of the pseudo 2D image compiled from background subtracted  $\chi$  dependent raw data is shown in Fig. 2. Due to the axial detector width of 8 mm, overlap of the measured diffraction patterns (Debye ring completeness) decreases with increasing  $2\theta$ , being larger than 100 % up to  $13.9^\circ 2\theta$  and still 66 % at  $21^\circ 2\theta$ .

### 2.4. Data treatment

LaB<sub>6</sub>-2D frames were used to calibrate the integration parameters (detector center, distance, tilt) of Dioptas v. 0.2.4 (Prescher & Prakapenka, 2015). All frames were subsequently integrated with Dioptas to .xy powder diffraction patterns with 2574 bins (equal to 0.0165° 2 $\theta$  per step at 0.2 m and 0.0098° 2 $\theta$  per step at 0.4 m). Intensity s.u.'s were obtained with SNBL Toolbox v. 0.5  $\sigma$ -scaler (Dyadkin, 2015). The same frames were also integrated with Fit2D (Hammersley *et al.*, 1996). Powder patterns matched those from Dioptas, except for slightly broader peaks, which most likely results from different data binning. Fit2D's maximum bin width is limited to the smallest width of the detector (1475 bins), whereas Dioptas uses twice the longest distance from the Debye ring center to the outmost pixel of the detector.

LaB<sub>6</sub>-1D raw data was post processed to an .xye diffraction pattern with a step size of 0.0036° 2 $\theta$  automatically after data acquisition.

All LaB<sub>6</sub> patterns were Pawley refined using TOPAS Academic v. 5 (Coelho, 2012) by applying a pseudo Voigt peak shape and the Caglioti equation (Caglioti *et al.*, 1958) to determine the instrumental resolution function. The Caglioti parameter U was set to 0 due to the restricted angular range (ca. 38° 2 $\theta$  before the DAC body shades the beam).

To obtain accurate L-Ser IV structure factors from the 144  $\chi$  dependent patterns of sample SerA- $\chi$ , the following procedure was applied (Fig. 3): (1) From each pattern, all peak intensities were extracted with Pawley refined unit cells of L-Ser I, II and IV (intensities, background, damped lattice parameters, peak broadening). The cell of L-Ser IV was previously indexed with TOPAS by using the singular value decomposition algorithm (Coelho, 2003), yielding an orthorhombic cell with  $P2_12_12_1$  as most probable space group with  $a = 16.09873(5)$  Å,  $b = 8.42773(6)$  Å,  $c = 8.04459(7)$  Å and  $V = 1091.46(2)$  Å<sup>3</sup>. (2) The obtained structure factors were cleaned from outliers  $\leq$  and  $\geq 3$  s.u.'s, and (3) a merged powder pattern was calculated using these average intensities. (4) This pattern was subsequently refined with a combined Rietveld (L-Ser I and II) Pawley (IV) approach in order to subtract L-Ser I and II intensities from overlapped peaks. Structural data of L-Ser I at 4.5 GPa were taken from Moggach *et al.* (2005) and those of L-Ser II at 6.3 GPa from Wood *et al.* (2008). From the final Pawley intensities, (5) a powder pattern was computed and used to solve the structure.

Steps 4 and 5 were also applied to the data of sample SerB-2D in order to compare the intensities and structures obtained from the  $\chi$  dependent data to those from the 2D data.

## 2.5. Structure Solution

By starting from a supercell of L-Ser I with a tripled  $a$ -axis, the structure of IV was solved by simulated annealing (Coelho, 2000). Refined parameters included a scale factor, translation along the unit cell axes (restricted to the three subcells in the  $P2_12_12_1$  supercell) and rotation of the three independent rigid-body molecules. The molecule geometry was taken from L-Ser I. Moreover,

rotation of the  $-\text{CH}_2\text{OH}$  and  $-\text{COO}$  group about their corresponding C–C axis (Fig. 4) was refined.  $B_{\text{iso}}$  values were set to 0. The resulting preliminary structure solution was used as starting model for periodic DFT calculations at B3LYP level of theory using full electron 6-31G(d,p) basis set with CRYSTAL09 (Dovesi *et al.*, 2009). The final structure ( $R_{\text{Bragg}} = 10.2\%$ ) was repeatedly obtained in several simulated annealing runs by using molecule geometries from the converged DFT calculation. This structure is discussed in detail in Fisch *et al.* (2015) and a CIF file can be downloaded from the Cambridge Structural Database (CCDC number 1412832).

### 3. Results and discussion

From the DAC containing  $\text{LaB}_6$ , the angular dependence of FWHM (Caglioti function parameters, Table 2) and peak shape of the two different setups was determined. The height-normalized 111 reflection of  $\text{LaB}_6$  of each dataset is shown in Fig. 5.

$\text{LaB}_6$  peaks measured with the 1D detector are much sharper than those obtained from the 2D data. The drastic difference in FWHM corresponds to a ratio of 10:1 with 0.2 m and still 6:1 for 0.4 m detector distance. The more than three times smaller pixel size of the curved Mythen II 1D detector compared to that of the 2D Pilatus 2M detector and almost twice the detector distance are the main reasons for the significantly higher resolution. Within the covered angular range, FWHM of the 1D data remains almost constant, however, slightly increasing peak broadening ( $V > 0$ , Table 2) due to the sample, equatorial and axial divergence and wavelength dispersion can be noticed and the pseudo Voigt peaks have ca. 27 % Lorentzian contribution. FWHM of the 2D data shows an opposite behavior ( $V < 0$ ), which is expected because the flat detector lies tangentially on the goniometer radius. X-rays diffracted at high diffraction angles impinge more tilted on the detector than at low angle and are therefore better resolved (Hinrichsen *et al.*, 2008). This effect is stronger than any other effect (e.g. sample broadening), and in sum, FWHM measured from  $\text{LaB}_6$  slightly decreases with increasing  $2\theta$ .

In practice, absolute intensities of the two setups are not comparable. Among the reasons are differences in primary beam intensities, detector response and post processing, absorption in air and irradiated sample volumes. Identification of the sharper peaks measured with the 1D detector is nonetheless more straightforward, also because they have a larger maximum height compared to broader ones with the same integrated area. Consequently, they are better resolved from their baseline.

The smaller FWHM is advantageous for the precision of peak position determination. As an example, the uncertainty of Pawley refined  $\text{LaB}_6$  lattice parameters is more than three times smaller in the 1D than in the 2D data. Obtained values are 4.1569160(16) Å and 4.1569165(59) Å for the 1D and the 2D data at 0.2 m, respectively, when refined from the same 10 reflections.

Sample SerB-1D was measured at one fixed  $\chi$  position, as well as with the 2D detector (sample SerB-2D) at 0.2 m detector distance. In the latter data, reflections of the pressure calibrant ruby sphere contribute to the diffraction pattern, whereas no ruby diffraction is observed in the 1D data. Enlarged regions of Pawley refined diffraction patterns obtained from both setups are shown in Fig. 6. The effect of the superior angular resolution of the 1D data becomes instantly noticeable: For example, the first two reflections of L-Ser II at  $9.21^\circ 2\theta$  and  $9.25^\circ 2\theta$  (A, Fig. 6) are well separated in the 1D data, whereas only a broad superimposed peak is observed in the 2D data. The same is true for the reflection of L-Ser I at  $10.53^\circ 2\theta$  and the one of IV at  $10.57^\circ 2\theta$  (B). Moreover, the intensity extraction of the sharper peaks of the 1D data is much more reliable: The value of the L-Ser IV peak at  $10.93^\circ 2\theta$  (C) is significant in the 1D data ( $> 5$  times its s.u.), whereas it is smaller than its s.u. in the 2D data. Moreover, even weak intensities are clearly identifiable, which results in a larger number of observed peaks in the SerA- $\chi$  data. Especially indexing of the relatively large unit cell of L-Ser IV would have been much more difficult from the SerB-2D data. From this powder pattern, 25 L-Ser IV reflections can be used for indexation, whereas 35 or 51 can be found in the SerB-1D or the SerA- $\chi$  data, respectively. When applying TOPAS' indexing method (Coelho, 2003) for all seven Bravais lattice types to the two 1D datasets, the three solutions with the highest goodness of fit yield the correct lattice parameters and space group, followed by other orthorhombic, tetragonal or monoclinic space groups containing a subset of the helical symmetry of the correct space group. In contrast, from the 25 reflections found in the SerB-2D data, only unit cells with two, four and eight times larger unit cells are among the best solutions (Fig. 7). Structural considerations enable to screen out non-chiral space groups for L-Serine, but such information may not always be available. From this data, the correct solution is also found but with a much smaller goodness of fit.

Extracted intensities of L-Ser IV can be better visualized by comparing simulated powder patterns based on normalized structure factors extracted from SerA- $\chi$  and SerB-2D, as shown in Fig. 8. The largest differences are above  $16^\circ 2\theta$ , which is the range in which the Debye ring coverage of the SerA- $\chi$  data has dropped below 100%. Moreover and unfortunately, the two samples do not contain the same relative amounts of L-Ser II and IV. The reason is that it is almost impossible to maintain exactly the same pressure increase rate during the compression of the DACs, which is a crucial factor for the formation of the different polymorphs, even when the same final pressure is reached (Fisch *et al.*, 2015). Rietveld quantitative phase analyses yielded 10 wt.-% L-Ser I, 12 wt.-% L-Ser II and 78 wt.-% L-Ser IV for sample SerA- $\chi$ , and 8 wt.-%, 38 wt.-% and 54 wt.-%, respectively for sample SerB-2D.

The severe peak overlap of polymorphs L-Ser I and II complicates the extraction of L-Ser IV structure factors. With the approach of a combined Rietveld (L-Ser I and II) and Pawley (IV) refinement, contributions of overlapped peaks of L-Ser I and II can be subtracted from the data, but this approach

is of course only as good as the structural models of L-Ser I at 4.5 GPa and II at 6.3 GPa fit the experimental data at the given pressure.

A preferred orientation correction on L-Ser II (210) improved the refinement significantly (March Dollase parameter of 0.885(5) for SerB-2D and 0.862(5) for SerA- $\chi$ ). In both datasets, L-Ser I also seems to suffer from preferred orientation, however, no correction was applied because the broad peaks are poorly defined and the chosen structural model is most likely slightly different from the one at the actual pressure. To discuss some of the differences in more detail, enlarged regions of the combined Rietveld (L-Ser I and II) and Pawley (IV) refined raw data are shown in Fig. 9.

Precise extraction of L-Ser IV intensities is clearly disturbed by the two strong ruby peaks (A, Fig. 9) contributing to the pattern of SerB-2D. In the 1D data, the corresponding L-Ser IV peaks are well separated from those of the other phases. Another discrepancy in the extracted intensities is found at  $17.9^\circ 2\theta$  (B). In the SerB-2D data, the measured peak is fully explained by a broad L-Ser II reflection, whereas in the sharper SerA- $\chi$  data, a significant amount of the peak area is also accounted for by L-Ser IV. The same feature can be observed at ca.  $16.1^\circ 2\theta$  (C). Even though the samples do not contain the same phase composition and, moreover, the coverage of Debye rings drops below 100 % in the SerA- $\chi$  data, the differences in intensity extraction are most likely due to the different peak resolutions. It is therefore unambiguous that the approach of intensity extraction from a multiphase sample is much more stable when applied to the SerA- $\chi$  data.

Refining the converged structure obtained from the SerA- $\chi$  against the SerB-2D data results in an increase of  $R_{\text{Bragg}}$  from 10.2 % to 15.6 %, and the structure becomes slightly distorted, compared to the one obtained from the DFT calculation. This is especially expressed in the rotation of molecule C and its  $-\text{CH}_2\text{OH}$  group torsion angle. A qualitative comparison of the final structure solutions and the DTF calculation is presented in Fig. 10. The root mean square of the distances between sites of the refined structures and the DFT geometry was compared using the IDEAL module implemented in WinGX v. 2013.3 (Farrugia, 2012). This value increases from 0.35 Å (SerA- $\chi$ ) to 0.55 Å (SerB-2D), indicating better agreement with the DFT geometry for the structure obtained from SerA- $\chi$ .

#### 4. Conclusions

We have presented the differences between HP powder data obtained from a 1D or 2D detector. In conventional setups, the former is suitable for obtaining accurate lattice parameters (e.g. for determination of equation of states), as well as highly-resolved peaks with small instrumental contribution to peak shape. The latter, in contrast, is clearly superior for obtaining accurate intensities. A method combining both qualities, i.e. HP P-XRD data with exceptionally high angular resolution and appropriate intensity statistics, has also been presented. The strategy, involving 1D measurements at several  $\chi$  positions, has been successfully applied to a challenging structure solution problem. This method will not replace conventional HP powder diffraction setups with 2D detectors, as the amount



of time and effort needed to perform the experiment is definitely of great disadvantage. However, it can be superior to conventional setups for special problems, such as indexing of large unit cells, dealing with multi-phase powder mixtures as well as for any investigation that requires high-resolution, e.g. microstructural studies at HP.

**Figure 1** Illustration of the powder diffractometer setup at the X04SA MS beamline. The detector (not fully shown) covers both negative and positive diffraction angles. For alignment purposes, the DAC can be moved along x and y inside the Eulerian cradle. During data collection,  $\chi$  is stationary and  $\omega$  oscillates.

**Figure 2** Left: The 2D data of sample SerB-2D is shown from  $0^\circ 2\theta$  to  $20^\circ 2\theta$ . The beam stop and gaps between detector modules are masked. Right: Pseudo 2D visualization of the background subtracted  $\chi$  dependent raw data of sample SerA- $\chi$ . Individual patterns from  $5^\circ 2\theta$  to  $20^\circ 2\theta$  are drawn as stripes on their corresponding  $\chi$  angle. Above  $13.9^\circ 2\theta$ , the sampling of the reciprocal space was not complete because of the chosen  $\chi$  steps. These regions are therefore missing.

**Figure 3** Data treatment procedure for extracting L-Ser IV structure factors from the diffraction patterns of samples SerA- $\chi$  and SerB-2D.

**Figure 4** L-Ser I superstructure starting model. During simulated annealing, rotation and translation (restricted to the subcells of the asymmetric unit, blue boxes) of each molecule, as well as rotation of the  $-\text{CH}_2\text{OH}$  and  $-\text{COO}$  groups around their corresponding C–C axis were refined.

**Figure 5** Raw data of the  $\text{LaB}_6$  111 reflection with normalized peak height. The secondary x-axis indicates half width at half maximum (HWHM) of the peaks.

**Figure 6** Enlarged regions of the Pawley refinements of samples SerB-2D (top) and SerB-1D (bottom). HKL ticks are shown below the diffraction patterns. The much better peak resolution of the 1D data becomes instantly visible. The details A, B and C are discussed in the text and the ruby peak in the 2D data is marked with R.

**Figure 7** Goodness of fit of indexing solutions for the 25 reflections extracted from the SerB-2D (red), the 35 found in the SerB-1D data (green) and the 51 from the SerA- $\chi$  data (blue) plotted against multiples of the unit cell volume. For each histogram, lattice parameters and space groups of the best solution are indicated. Structural considerations enable to screen out non-chiral space groups for L-Serine, but this plot gives a more comprehensive overview of the problems associated with each dataset.

**Figure 8** Calculated powder patterns based on the extracted structure factors of the SerA- $\chi$  and the SerB-2D samples. The enlarged difference curve represents  $I_{\text{SerA-}\chi} - I_{\text{SerB-2D}}$ .

**Figure 9** Enlarged regions of the combined Rietveld/Pawley refined patterns of sample SerB-2D (top) and SerA- $\chi$  (bottom). The details A, B and C are discussed in the text and the ruby peaks in the 2D data are marked with R.

**Figure 10** Asymmetric units of the final structures obtained from L-Ser IV intensities extracted from samples SerB-2D (left) and SerA- $\chi$  (right). The DFT calculated structure (green) is shown as reference. The RMS distance to the DFT structure increases from 0.35 Å (SerA- $\chi$ ) to 0.55 Å (SerB-2D).

**Table 1** Overview of the samples and measurement types.

Sample name	Measurement type	Beamline	Compound	Pressure
SerA- $\chi$	Pseudo 2D (360° $\chi$ )	MS	L-Serine	6.74(2) GPa
SerB-1D	1D (one $\chi$ position)	MS	L-Serine	6.82(2) GPa
SerB-2D	2D (0.2 m det. dist.)	PXIII	L-Serine	6.82(2) GPa
LaB <sub>6</sub> -1D	1D (one $\chi$ position)	MS	LaB <sub>6</sub>	0.0001 GPa
LaB <sub>6</sub> -2D	2D (0.2 and 0.4 m)	PXIII	LaB <sub>6</sub>	0.0001 GPa

**Table 2** Caglioti and peak shape parameters refined from the 1D and 2D LaB<sub>6</sub> datasets.

Dataset	W	V	U	$\eta^*$
LaB <sub>6</sub> -2D (0.2 m)	0.00723(6)	-0.0031(3)	0	0
LaB <sub>6</sub> -2D (0.4 m)	0.00357(6)	-0.0010(3)	0	0

---

LaB <sub>6</sub> -1D	0.000050(1)	0.000173(6)	0	0.271(2)
----------------------	-------------	-------------	---	----------

---

\* Pseudo Voigt mixing parameter  $\eta = 0$  for Gaussian profiles.

**Acknowledgements** We acknowledge support from the Swiss National Science Foundation (Project 144534) and Dr. Vincent Olieric for the kind and straightforward access to the PXIII beamline at SLS. M. F. thanks Dr. Alan Coelho for fruitful discussions.

## References

- Ashiotis, G., Deschildre, A., Nawaz, Z., Wright, J. P., Karkoulis, D., Picca, F. E. & Kieffer, J. (2015). *J. Appl. Cryst.* **48**, 510–519.
- Benecke, G., Wagermaier, C. & Li C. (2014). *J. Appl. Cryst.* **47**, 1797–1803.
- Bergamaschi, A., Cervellino, A., Dinapoli, R., Gozzo, F., Heinrich, B., Johnson, I., Kraft, P., Mozzanica, A., Schmitt, B. & Shi, X. T. (2010). *J. Synchrotron Rad.* **17**, 653–668.
- Boldyreva, E. V., Sowa, H., Seryotkin, Yu. V., Drebuschak, T. N., Ahsbahs, H., Chernyshev, V., Dmitriev, V. (2006). *Chem. Phys. Lett.* **429**, 474–478.
- Boldyreva, E. V. (2007). *Cryst. Growth. Des.* **7**, 1662–1668.
- Broennimann, Ch., Eikenberry, E. F., Henrich, B., Horisberger, R., Huelsen, G., Pohl, E., Schmitt, B., Schulze-Briese, C., Suzuki, M., Tomizaki, T., Toyokawa, H. & Wagner, A. (2006). *J. Synchrotron Rad.* **13**, 120–130.
- Caglioti, G., Paoletti, A. & Ricci, F. P. (1958). *Nucl. Instrum. Meth.* **3**, 223–228.
- Coelho, A. A. (2003). *J. Appl. Cryst.* **36**, 86–95.
- Coelho, A. A. (2000). *J. Appl. Cryst.* **33**, 899–908.
- Coelho, A. A. (2012). TOPAS-Academic v. 5, Coelho Software, Brisbane, Australia.
- Dovesi, R., Saunders, V. R., Roetti, C., Orlando, R., Zicovich-Wilson, C. M., Pascale, F., Civalleri, B., Doll, K., Harrison, N. M., Bush, I. J., D'Arco, P. & Llunell, M. (2009). CRYSTAL09, University of Torino, Torino, Italy.
- Dyadkin, V. (2015). SNBL Toolbox v. 0.5, Swiss-Norwegian Beamlines, European Synchrotron Radiation Facility, Grenoble, France.
- Evans, W. J.; Yoo, C. S.; Lee, G. W.; Cynn, H.; Lipp, M. J.; Visbeck, K. (2007). *Rev. Sci. Instrum.* **78**, 073904.
- Farrugia, L. J. (2012). *J. Appl. Cryst.* **45**, 849–854.
- Fisch, M., Lanza, A., Boldyreva E. V., Macchi, P. & Casati, N. (2015). *J. Phys. Chem. C*. DOI: 10.1021/acs.jpcc.5b05838.
- Hammersley, A. P., Svensson, S. O., Hanfland, M., Fitch, A. N. & Häusermann, D. (1996). *High Press. Res.* **14**, 235–248.
- Heiney, P. A. (2005). *Comm. Powder Diffr. Newsletter* **32**, 3-11.

- Hinrichsen, B., Dinnebier, R. E. & Jansen, M. (2008). Powder Diffraction Theory and Practice, edited by R. E. Dinnebier & S. J. Billinge, pp. 414–438. Cambridge: RCS.
- Ilavsky, J. (2012). *J. Appl. Cryst.* **45**, 324–328.
- Jamieson, J. C., Lawson, A. W. & Nachtrieb, N. D. (1959). *Rev. Sci. Instrum.* **30**, 1016–1019.
- Lanza, A., Fiolka, C., Fisch, M., Casati, N., Skoulatos M., Rüegg C., Krämer K. W. & Macchi P. (2014). *Chem. Commun.* **50**, 14504–14507.
- Moggach, S. A., Allan, D. R., Morrison, C. A., Parsons, S. & Sawyer, L. (2005) *Acta Cryst.* **B61**, 58–68.
- Perillat, J. P. (2008). *Min. Mag.* **72**, 683–695.
- Prescher, C. & Prakapenka, V. B. (2015). *High Pressure Res.* DOI:10.1080/08957959.2015.1059835.
- Rodriguez-Navarro, A. B. (2006). *J. Appl. Cryst.* **39**, 905–909.
- Shekar, N. V. C. & Rajan, K. G. (2001). *Bull. Mater. Sci.* **24**, 1–21.
- Smith, J. S. & Desgreines, S. (2008). *J. Synchrotron Rad.* **16**, 83–96.
- Syassen, K. (2008). *High Press. Res.* **28**, 75–126.
- Toby, B. H. & Von Dreele, R. B. (2013). *J. Appl. Cryst.* **46**, 544–549.
- Tschauner, O., McClure, J. & Nicol, M. (2005). *J. Synchrotron Rad.* **12**, 626–631.
- Velisavljevic, N.; Sinogeikin, S.; Saavedra, R.; Chellappa, R. S.; Rothkirch, A.; Dattelbaum, D. M.; Konopkova, Z.; Liermann, H. P.; Bishop, M.; Tsoi, G.M.; *et al.* (2014). *J. Phys. Conf. Ser.* **500**, 032020.
- Waltersperger, S., Olieric, V., Pradervand, C., Gletting, W., Salathe, M., Fuchs, M. R., Curtin, A., Wang, X., Ebner, S., Panepucci, E., Weinert, T., Schulze-Briese, C., & Wang, M. (2015), *J. Synchrotron Rad.* **22**, 895–900.
- Willmott, P. R., Meister, D., Leake, S. J., Lange, M., Bergamaschi, A., Böge, M., Calvi, M., Cancellieri, C., Casati, N., Cervellino, A., Chen, Q., David, C., Flechsig, U., Gozzo, F., Henrich, B., Jäggi-Spielmann, S., Jakob, B., Kalichava, I., Karvinen, P., Krempasky, J., Lüdeke, A., Lüscher, R., Maag, S., Quitmann, C., Reinle-Schmitt, M. L., Schmidt, T., Schmitt, B., Streun, A., Vartiainen, I., Vitins, M., Wang, X. & Wulschleger, R. (2013). *J. Synchrotron Rad.* **20**, 667–682.
- Wood, P. A., Francis, D., Marshall, W. G., Moggach, S. A., Parsons, S., Pidcock, E. & Rohl, A. L. (2008). *CrystEngComm* **10**, 1154–1166.