

RESEARCH

Open Access

500 ml of blood loss does not decrease non-invasive tissue oxygen saturation (StO₂) as measured by near infrared spectroscopy - A hypothesis generating pilot study in healthy adult women

Victor Jeger¹, Stephan M Jakob², Stefano Fontana³, Martin Wolf⁴, Heinz Zimmermann¹ and Aristomenis K Exadaktylos*¹

Abstract

Background: The goal when resuscitating trauma patients is to achieve adequate tissue perfusion. One parameter of tissue perfusion is tissue oxygen saturation (StO₂), as measured by near infrared spectroscopy. Using a commercially available device, we investigated whether clinically relevant blood loss of 500 ml in healthy volunteers can be detected by changes in StO₂ after a standardized ischemic event.

Methods: We performed occlusion of the brachial artery for 3 minutes in 20 healthy female blood donors before and after blood donation. StO₂ and total oxygenated tissue hemoglobin (O₂Hb) were measured continuously at the thenar eminence. 10 healthy volunteers were assessed in the same way, to examine whether repeated vascular occlusion without blood donation exhibits time dependent effects.

Results: Blood donation caused a substantial decrease in systolic blood pressure, but did not affect resting StO₂ and O₂Hb values. No changes were measured in the blood donor group in the reaction to the vascular occlusion test, but in the control group there was an increase in the O₂Hb rate of recovery during the reperfusion phase.

Conclusion: StO₂ measured at the thenar eminence seems to be insensitive to blood loss of 500 ml in this setting. Probably blood loss greater than this might lead to detectable changes guiding the treating physician. The exact cut off for detectable changes and the time effect on repeated vascular occlusion tests should be explored further. Until now no such data exist.

Introduction

The goal of resuscitation is to achieve adequate tissue perfusion. The patient's history, ongoing bleeding and clinical signs of hypovolemia are the first trigger for volume substitution. Tissue perfusion is conventionally monitored using arterial base deficit and serum lactate, but these laboratory parameters are not always immediately available at the point-of-care. Additionally, relevant

blood loss may not be detected by conventional hemodynamic monitoring [1].

Therefore, in the trauma setting, there is growing interest in non-invasive techniques measuring tissue perfusion. Near infrared spectroscopy (NIRS) has become a widely used method for tissue hemoglobin oxygen saturation (StO₂) measurement in muscle and has been validated in animals [2-4] and in humans [5-11].

NIRS uses light with a wavelength of 650-950 nm and easily crosses biological tissues. Near infrared light is mainly absorbed by hemoglobin and only monitors vessels with a diameter <1 mm, because the blood levels in

* Correspondence: exadaktylos@exadaktylos.ch

¹ Department of Emergency Medicine, University and University Hospital of Berne (Inselspital), Berne, Switzerland

Full list of author information is available at the end of the article

larger vessels are too high to reflect enough light [12]. In this setting, NIRS measurements primarily indicate the venous oxyhemoglobin concentration, as only 20% of blood volume is arterial and the NIRS device does not differentiate between systole and diastole [13].

NIRS can be used to observe the microcirculatory reactivity of peripheral tissue after a standardized ischemic event. It has been shown that StO_2 recovery after transient occlusion of the brachial artery differs in hemodynamically unstable trauma patients in comparison to healthy volunteers [14]. The StO_2 response after a vascular occlusion test is an example of functional hemodynamic monitoring, in which the response of the specific system to a pre-determined stress is the monitored variable. After induction of an artificial ischemic stress, local metabolic demand and reperfusion reserve can be assessed from the changes in parameters [14].

The aim of this study was to investigate the impact of a controlled and clinically relevant blood loss (500 ml) on resting StO_2 values, as well as microcirculatory reactivity triggered by a vascular occlusion test in healthy blood donors. 500 ml of blood donation correspond to a loss of 10% of the whole blood volume. 10% of blood loss is considered to be the cut off for clinical relevant blood loss according to the current literature and may become relevant if pre-existing morbidity exists. For example, postpartum hemorrhage, defined as blood loss more than 500 ml after a vaginal delivery, is a major cause of maternal morbidity and mortality [15]. Other authors have been able to detect 500 ml of blood loss with NIRS in various settings [16,17].

We have decided to investigate the 500 ml blood loss because of the controlled and safe study setting and because higher blood volume losses are anyway related with detectable changes in physiological parameters (e.g. blood pressure, pulse rate). For the clinician, "borderline" patients without obvious signs of blood loss present the greater challenge.

We hypothesized that NIRS after a vascular occlusion test would be able to detect this amount of blood loss, because of the industries promise of "earliest" detection of volume depletion.

Methods

Volunteers

We investigated 20 healthy female blood donors and 10 healthy female medical students as control subjects. The study was restricted to women, as they have lower body mass index (BMI) than men and are therefore more prone to show slight symptoms of hypovolemia after blood donation.

Blood donors had a median age of 30.5 y (range 19 - 62). Their BMI was 21 kg/m² median (range 18 - 23.5). The

median blood volume is estimated as 70 ml/kg body weight or 4060 ml (range 3500 - 4550). The median capillary hemoglobin was 139 g/l (range 127 - 157). This was measured once immediately after arrival at the blood donation service.

In the control group, the median age was 23 y (range 22 - 25). The median BMI was 21 kg/m² (range 19 - 23). The median blood volume was 4165 ml (range 3500 - 4480).

Protocol

Blood donors and control subjects were in a zero degree supine position during the whole test. In both groups, there was a resting period of 5 minutes before applying the probe and performing the vascular occlusion test measurement. We measured blood pressure manually before every vascular occlusion test, using the same sphygmomanometer cuff as when occluding the artery. Blood donation started after a recovery period of 5 minutes. The second vascular occlusion test in the blood donor group started immediately after the end of blood donation.

In the control group, measurements were performed in exactly the same way as in the blood donor group. In order to detect a possible time effect from the vascular occlusion tests, a 19-minute rest period was maintained between the two vascular occlusion tests of the control group. (Fig. 1) These 19 minutes are a median (range 14 - 22 minutes) measured in the blood donor group between the end of the first and the start of the second vascular occlusion test. This period included the recovery from the first vascular occlusion, the preparation for phlebotomy and the actual blood donation, which lasted 11 minutes (range 9 - 13).

As this was an observational study aimed at controlling the quality of a freely available device, and because no additional data were collected from the volunteers, no formal ethical approval was necessary, according to our local ethical committee. Additionally, the vascular occlusion using a cuff pressure of 50 mmHg above systolic blood pressure, correlates to cuff pressures reached in non invasive (Riva Rocci) blood pressure measurement. Because no data on ethical considerations related to such kind of setting have been available, AKE and VJ have acted as volunteers. The discomfort has been within the usual limits of blood pressure measurement and was later very well tolerated by the volunteers and did not cause any discomfort.

Measurements

We measured StO_2 at the thenar eminence (opposite to the blood donating arm) with a commercially available tissue spectrometer (InSpectra® Model 650, Hutchinson Technology, Hutchinson, Minn. US), which is promoted as being able to detect blood loss in an emergency setting




Figure 1 Time axis showing course of events for both groups.

[18]. The method has been previously validated [19]. The probe spacing was 15 mm.

The device measures and records StO_2 and tissue hemoglobin index (THI) every two seconds. THI is a quantitative estimation of tissue hemoglobin concentration [20]. From these values, oxygenated hemoglobin (O_2Hb) was obtained by multiplying StO_2 and THI, because $StO_2 = \frac{O_2Hb}{O_2Hb+HHb}$ [19]. Accordingly, the amount of deoxygenated hemoglobin (HHb) equals the residual hemoglobin: $THI - O_2Hb = HHb$. We defined baseline variables before starting a vascular occlusion test as the median over 2 minutes for StO_2 , O_2Hb , and HHb.

A vascular occlusion test was performed by inflating the sphygmomanometer cuff for 3 minutes with a pressure of 50 mmHg over systolic BP as described in detail by Creteur et al. [12]. The rate of decrease was calculated from the linear part of the decrease (first 25%) after cuff inflation [14]. The rate of recovery of StO_2 and O_2Hb was defined over a mean of 14 seconds (8 data points) following the nadir StO_2 and O_2Hb value at the end of ischemia. Delta StO_2 and delta O_2Hb are defined as the difference between the maximal value after reperfusion and the baseline [12]. (Fig. 2, Fig. 3)

Data processing

Prism 5 (GraphPad Software, Inc., San Diego, CA, USA) was used to calculate the statistics. Non-parametric tests were used (Wilcoxon rank sum test for within-group differences, Mann-Whitney U-test for between-group differences). Data is presented as medians and interquartile ranges. P-values have been interpreted in an exploratory fashion.

Results

There was a markedly decrease in systolic blood pressure in both groups before and after blood donation. Systolic






Figure 2 Typical StO_2 -trace of a blood donor. Whole length of recorded trace showing all events: 1st vascular occlusion test, recovery period, blood donation and 2nd vascular occlusion test. Details of a vascular occlusion test are described in figure 3.



blood pressure decreased from 118 mmHg median (interquartile range: 118 - 123) to 110 mmHg (106 - 115) ($p < 0.001$) in the donor group and from 112 mmHg (103 - 118) to 106 mmHg (100 - 111) ($p = 0.02$) in the control group. Systolic blood pressure decreased markedly more in the donor group than in the control group ($p = 0.02$). (Fig. 4)



For variables measured with NIRS, the differences from the first to the second vascular occlusion test did not differ between the two groups. The only exception was the differences in O_2Hb 2 min median, which was slightly larger in the control group -0.4 ($-1.0 - -0.3$) than in the blood donor group -0.1 ($-0.4 - 0.3$) ($p = 0.03$). Therefore the two groups are considered to be similar and comparable.

StO_2 and O_2Hb 2 min median did not change substantially in the blood donor group before and after blood donation, but decreased in the control group after the rest period: StO_2 2 min median decreased from 82.0% (79.5 - 85.0) to 80.5% (76.8 - 83.8) ($p = 0.02$), O_2Hb 2 min median decreased from 10.9 (9.6 - 12.4) to 10.2 (9.4 - 11.8) ($p = 0.03$). HHb 2 min median did not change in the blood donor group, but increased in the control group markedly from 2.35 (2.08 - 2.56) to 2.45 (2.20 - 2.92) ($p = 0.01$) (Table 1).

Variables measured during the vascular occlusion test did not change in the blood donor group: The rate of decrease, rate of recovery and delta for StO_2 and O_2Hb did not show any substantial difference before and after blood donation. In the control group, StO_2 variables did not change markedly - in contrast to the O_2Hb values. O_2Hb rate of decrease was less steep after the rest period (before: -0.039 ($-0.051 - -0.026$); after: -0.031 ($-0.046 - -0.020$); $p < 0.01$), whereas O_2Hb rate of recovery (before: 0.49 ($0.39 - 0.63$); after: 0.54 ($0.39 - 0.69$); $p = 0.05$) and O_2Hb delta (before: 3.88 ($2.43 - 4.20$); after: 4.0 ($3.32 - 5.72$); $p = 0.01$) increased after the rest period (Table 2).

Discussion

The principal finding of this study is, that although systolic blood pressure decreased substantially, blood loss had no effect on superficial thenar StO_2 before and after blood donation measured by NIRS. However other experimental NIRS devices have been successfully used to monitor blood loss after blood donation. Torella et al. showed that even 2% blood volume loss reduced peripheral hemoglobin oxygen saturation (PsO_2 , calf) significantly [16]. Soller et al. simulated hypovolemia using progressive lower body negative pressure and showed that hypovolemia corresponding to 400 - 500 ml of blood loss resulted in a decrease in forearm muscle oxygen saturation [17]. The same group could also show, in the same setting, that deep muscle oxygen saturation is more sensitive in detecting hypovolemia (experimental NIRS device) than superficial thenar StO_2 (Hutchinson) [21]. Soller et al. discussed, that the location of the NIRS sensor as well as the tissue depth might be responsible for variations in results, although the technique used by the two devices is in principle the same. We did use exactly

Table 1: NIRS variables and systolic blood pressure before each vascular occlusion test.

Blood donors	Before blood donation		After blood donation		two-tailed p-value
	median	IQR	median	IQR	
StO ₂	79.0	74.5 - 83.5	78.5	75.0 - 83.0	0.34
O ₂ Hb	10.7	9.3 - 11.9	10.8	8.9 - 11.7	0.43
deoxy Hb	2.62	2.39 - 3.24	2.70	2.34 - 4.39	0.30
SBP	118	118 - 123	110	106 - 115	<0.001*

Control	Before rest period		After rest period (19 minutes)		two-tailed p-value
	median	IQR	median	IQR	
StO ₂	82.0	79.5 - 85.0	80.5	76.8 - 83.8	<0.05*
O ₂ Hb	10.9	9.6 - 12.4	10.2	9.4 - 11.8	<0.05*
deoxy Hb	2.35	2.08 - 2.56	2.45	2.20 - 2.92	<0.01*
SBP	112	103 - 118	106	100 - 111	<0.05*

IQR: interquartile range, *: p < 0.05

Table 2: NIRS variables during the vascular occlusion tests.

Blood donors	Before blood donation		After blood donation		two-tailed p-value
	median	IQR	median	IQR	
StO ₂ rate of decrease(%/sec)	-0.19	-0.21 - -0.16	-0.19	-0.23 - -0.15	0.78
StO ₂ rate of recovery (%/sec)	3.04	2.57 - 3.45	3.29	2.79 - 3.48	0.25
StO ₂ delta (%)	14.0	12.0 - 15.8	13.5	12.0 - 18.0	0.36
O ₂ Hb rate of decrease (/sec)	-0.033	-0.043 - -0.026	-0.027	-0.055 - -0.021	0.98
O ₂ Hb rate of recovery (/sec)	0.59	0.50 - 0.72	0.63	0.48 - 0.74	0.59
O ₂ Hb delta	4.33	3.47 - 4.97	4.76	4.24 - 5.42	0.06

Control	Before rest period		After rest period (19 minutes)		two-tailed p-value
	median	IQR	median	IQR	
StO ₂ rate of decrease (%/sec)	-0.17	-0.19 - -0.15	-0.15	-0.20 - -0.12	0.43
StO ₂ rate of recovery (%/sec)	2.64	2.00 - 2.88	2.43	2.00 - 3.47	0.29
StO ₂ delta (%)	12.5	11.3 - 13.3	13.5	11.5 - 15.25	0.10
O ₂ Hb rate of decrease (/sec)	-0.039	-0.051 - -0.026	-0.031	-0.046 - -0.020	<0.01*
O ₂ Hb rate of recovery (/sec)	0.49	0.39 - 0.63	0.54	0.39 - 0.69	<0.05*
O ₂ Hb delta	3.88	2.43 - 4.20	4.00	3.32 - 5.72	<0.01*

IQR: interquartile range, *: p < 0.05

the same superficial thenar StO₂ probe from Hutchinson as Soller in her study, which may explain why we did not detect any changes in NIRS parameters following blood loss.

A second finding of this study is that there might be an intrinsic effect of repeated vascular occlusion tests on O₂Hb. This hypothesis is based on the increase in delta O₂Hb and O₂Hb rate of recovery and a decrease in oxygen consumption (O₂Hb rate of decrease), which was measured in the control group but not in the blood donor group. This could be explained if there is a time effect of repeated vascular occlusion tests, designated as ischemic preconditioning [22]. This improves the ischemic tolerance of various tissues, including skeletal muscle, after one or more brief periods of ischemia, each followed by a short reperfusion phase. The underlying mechanisms are still not clear. One reason may be the reduction in the resting rate of oxygen consumption [23]. Although blood loss does not result in different StO₂ or O₂Hb variables after the vascular occlusion test, it seems to mask the effect of ischemic preconditioning in the blood donor group.

Limitations

One limitation with our paper is the assumption that withdrawal of 500 ml of blood in a healthy volunteer represents a clinically relevant blood loss. Anyhow 10% of circulating blood volume can not be neglected.

In addition we did not supply data on base excess and/or lactate, the two parameters which are conventionally used to estimate tissue oxygenation. This would have helped to indicate whether a 500 ml blood loss was indeed relevant, but was not feasible in the giving setting of volunteer blood donation. Therefore, it is unclear as to whether the device failed to identify tissue ischemia completely, which is unlikely as the values decreased during vascular occlusion, or more likely, that no ischemia developed during blood donation.

Conclusion

StO₂ measured at the thenar eminence seems to be insensitive to blood loss of 500 ml in this setting. Probably blood loss greater than this might lead to detectable changes guiding the treating physician. The exact cut off for detectable changes and the time effect on repeated vascular occlusion tests should be explored further. Until now no such data exist. Finally, our findings suggest that StO₂ measured by NIRS is not able to detect acute blood loss of 500 ml and therefore, it might not be useful to apply this device in the acute assessment of a bleeding trauma patient. However, the device could help to identify hemodynamic worsening during the ICU stay but at

the moment not enough evidence exists to support this hypothesis.

Competing interests

The authors declare that they have no competing interests.

Authors' contributions

VJ, SF and AE conducted the study. All authors contributed to the design of the study, the analysis of the data and the writing of the manuscript. All authors have seen the original study data, reviewed the analysis of the data, and approved the final manuscript. VJ is the author responsible for archiving the study files.

Acknowledgements

Funding: The present study has been performed without financial support from outside the department of Emergency Medicine.

Previous exhibition: This report has been previously presented, in abstract form, at the ESICM annual congress 2008.

Author Details

¹Department of Emergency Medicine, University and University Hospital of Berne (Inselspital), Berne, Switzerland, ²Department of Intensive Care Medicine, University and University Hospital of Berne (Inselspital), Berne, Switzerland, ³Blood Donation Service SRK, Berne, Switzerland and ⁴Biomedical Optics Research Laboratory, Clinic of Neonatology, University Hospital Zurich, Switzerland

Received: 7 December 2009 Accepted: 13 May 2010

Published: 13 May 2010

References

1. Hamilton-Davies C, Mythen MG, Salmon JB, Jacobson D, Shukla A, Webb AR: **Comparison of commonly used clinical indicators of hypovolaemia with gastrointestinal tonometry.** *Intensive Care Med* 1997, **23**:276-281.
2. Cohn SM, Varela JE, Giannotti G, Dolich MO, Brown M, Feinstein A, McKenney MG, Spalding P: **Splanchnic perfusion evaluation during hemorrhage and resuscitation with gastric near-infrared spectroscopy.** *J Trauma* 2001, **50**:629-634. discussion 634-625
3. Varela JE, Cohn SM, Giannotti GD, Dolich MO, Ramon H, Wiseberg JA, McKenney M: **Near-infrared spectroscopy reflects changes in mesenteric and systemic perfusion during abdominal compartment syndrome.** *Surgery* 2001, **129**:363-370.
4. Chaisson NF, Kirschner RA, Deyo DJ, Lopez JA, Prough DS, Kramer GC: **Near-infrared spectroscopy-guided closed-loop resuscitation of hemorrhage.** *J Trauma* 2003, **54**:S183-192.
5. Cohn SM, Nathens AB, Moore FA, Rhee P, Puyana JC, Moore EE, Beilman GJ: **Tissue oxygen saturation predicts the development of organ dysfunction during traumatic shock resuscitation.** *J Trauma* 2007, **62**:44-54. discussion 54-45
6. Taylor DE, Simonson SG: **Use of near-infrared spectroscopy to monitor tissue oxygenation.** *New Horiz* 1996, **4**:420-425.
7. Giannotti G, Cohn SM, Brown M, Varela JE, McKenney MG, Wiseberg JA: **Utility of near-infrared spectroscopy in the diagnosis of lower extremity compartment syndrome.** *J Trauma* 2000, **48**:396-399. discussion 399-401
8. Boushel R, Piantadosi CA: **Near-infrared spectroscopy for monitoring muscle oxygenation.** *Acta Physiol Scand* 2000, **168**:615-622.
9. van Beekvelt MC, Borghuis MS, van Engelen BG, Wevers RA, Colier WN: **Adipose tissue thickness affects in vivo quantitative near-IR spectroscopy in human skeletal muscle.** *Clin Sci (Lond)* 2001, **101**:21-28.
10. Rothoerl RD, Faltermeier R, Burger R, Woertgen C, Brawanski A: **Dynamic correlation between tissue PO₂ and near infrared spectroscopy.** *Acta Neurochir Suppl* 2002, **81**:311-313.
11. Pichler G, Urlesberger B, Jirak P, Zotter H, Muller W: **Forearm oxygen consumption and forearm blood flow in healthy children and adolescents measured by near infrared spectroscopy.** *J Physiol Sci* 2006, **56**:191-194.
12. Creteur J, Carollo T, Soldati G, Buchele G, De Backer D, Vincent JL: **The prognostic value of muscle StO₂ in septic patients.** *Intensive Care Med* 2007, **33**:1549-1556.

13. Crookes BA, Cohn SM, Bloch S, Amortegui J, Manning R, Li P, Proctor MS, Hallal A, Blackbourne LH, Benjamin R, et al.: **Can near-infrared spectroscopy identify the severity of shock in trauma patients?** *J Trauma* 2005, **58**:806-813. discussion 813-806
14. Gomez H, Torres A, Polanco P, Kim HK, Zenker S, Puyana JC, Pinsky MR: **Use of non-invasive NIRS during a vascular occlusion test to assess dynamic tissue O(2) saturation response.** *Intensive Care Med* 2008, **34**:1600-1607.
15. Toledo P, McCarthy RJ, Hewlett BJ, Fitzgerald PC, Wong CA: **The accuracy of blood loss estimation after simulated vaginal delivery.** *Anesth Analg* 2007, **105**:1736-1740. table of contents
16. Torella F, Cowley RD, Thorniley MS, McCollum CN: **Regional tissue oxygenation during hemorrhage: can near infrared spectroscopy be used to monitor blood loss?** *Shock* 2002, **18**:440-444.
17. Soller BR, Yang Y, Soyemi OO, Ryan KL, Rickards CA, Walz JM, Heard SO, Convertino VA: **Noninvasively determined muscle oxygen saturation is an early indicator of central hypovolemia in humans.** *J Appl Physiol* 2008, **104**:475-481.
18. Smith J, Bricker S, Putnam B: **Tissue oxygen saturation predicts the need for early blood transfusion in trauma patients.** *Am Surg* 2008, **74**:1006-1011.
19. Myers DE, Anderson LD, Seifert RP, Ortner JP, Cooper CE, Beilman GJ, Mowlem JD: **Noninvasive method for measuring local hemoglobin oxygen saturation in tissue using wide gap second derivative near-infrared spectroscopy.** *J Biomed Opt* 2005, **10**:034017.
20. Skarda DE, Mulier KE, Myers DE, Taylor JH, Beilman GJ: **Dynamic near-infrared spectroscopy measurements in patients with severe sepsis.** *Shock* 2007, **27**:348-353.
21. Soller BR, Ryan KL, Rickards CA, Cooke WH, Yang Y, Soyemi OO, Crookes BA, Heard SO, Convertino VA: **Oxygen saturation determined from deep muscle, not thenar tissue, is an early indicator of central hypovolemia in humans.** *Crit Care Med* 2008, **36**:176-182.
22. Gurke L, Mattei A, Chaloupka K, Marx A, Sutter PM, Stierli P, Harder F, Heberer M: **Mechanisms of ischemic preconditioning in skeletal muscle.** *J Surg Res* 2000, **94**:18-27.
23. Sen A, Martin D, Grocott M, Montgomery H: **Changes in thenar eminence tissue oxygen saturation measured using near-infrared spectroscopy suggest ischaemic preconditioning in a repeated arterial occlusion forearm ischaemia model.** *Critical Care* 2008, **12**(Suppl 2):P70. (Abstract)

doi: 10.1186/1752-2897-4-5

Cite this article as: Jeger et al., 500 ml of blood loss does not decrease non-invasive tissue oxygen saturation (StO2) as measured by near infrared spectroscopy - A hypothesis generating pilot study in healthy adult women *Journal of Trauma Management & Outcomes* 2010, **4**:5

Submit your next manuscript to BioMed Central and take full advantage of:

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at
www.biomedcentral.com/submit

